


LB136027

7471B

INSTRUMENT ID: CV1

 μ Abs.:

56593



Linear

A= 0.0000e+000

B= 1.7517e-004

C= -3.2719e-002

Rho= 0.9997971

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.029	-0.029	20	0.000	20					4/11
0.2	0.200	0.175	-0.025	1187	0.0 %	1187					1
2.5	2.500	2.493	-0.007	14420	0.0 %	14420					13
5.0	5.000	5.067	0.067	29116	0.0 %	29116					0
7.5	7.500	7.613	0.113	43649	0.0 %	43649					2
10.0	10.000	9.880	-0.120	56593	0.0 %	56593					1

LB136027 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Jun 2025 16:29:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	7471B	PPB	05 Jun 2025 16:46:18	S	Std
0.2 - 1	50.2	1187	-	0.2000	7471B	PPB	05 Jun 2025 16:48:35	S	Std
2.5 - 1	52.5	14420	-	2.5000	7471B	PPB	05 Jun 2025 16:50:52	S	Std
5.0 - 1	55	29116	-	5.0000	7471B	PPB	05 Jun 2025 16:53:09	S	Std
7.5 - 1	57.5	43649	-	7.5000	7471B	PPB	05 Jun 2025 16:55:28	S	Std
10.0 - 1	59.5	56593	-	10.0000	7471B	PPB	05 Jun 2025 17:02:55	S	Std
ICV28 - 1	ICV28	22091	3.8369	-7471B	PPB	05 Jun 2025 17:06:26	U	SMPL	
ICB28 - 1	ICB28	-154	-0.0597	-7471B	PPB	05 Jun 2025 17:08:40	U	SMPL	
CCV04 - 1	CCV04	29086	5.0621	-7471B	PPB	05 Jun 2025 17:10:58	U	SMPL	
CCB04 - 1	CCB04	-186	-0.0653	-7471B	PPB	05 Jun 2025 17:13:13	U	SMPL	
CRA - 1	CRA	1243	0.1850	-7471B	PPB	05 Jun 2025 17:15:31	U	SMPL	
ChkStd - 1	ChkStd	39723	6.9254	-7471B	PPB	05 Jun 2025 17:20:01	U	SMPL	
PB168306BL - 1	PBS	310	0.0216	-7471B	PPB	05 Jun 2025 17:22:17	U	SMPL	
PB168306BS - 1	LCSS	20878	3.6244	-7471B	PPB	05 Jun 2025 17:24:39	U	SMPL	
Q2206-01 - 1	TP-1	1876	0.2959	-7471B	PPB	05 Jun 2025 17:37:50	U	SMPL	
Q2207-01 - 1	BU-703-COMP-01	2042	0.3250	-7471B	PPB	05 Jun 2025 17:40:06	U	SMPL	
Q2207-10 - 1	BU-703-COMP-02	2921	0.4789	-7471B	PPB	05 Jun 2025 17:42:24	U	SMPL	
Q2207-19 - 1	BU-703-COMP-03	3940	0.6574	-7471B	PPB	05 Jun 2025 17:44:46	U	SMPL	
Q2207-28 - 1	BU-703-COMP-04	1911	0.3020	-7471B	PPB	05 Jun 2025 17:47:11	U	SMPL	
CCV05 - 1	CCV05	29155	5.0742	-7471B	PPB	05 Jun 2025 17:49:29	U	SMPL	
CCB05 - 1	CCB05	-60	-0.0432	-7471B	PPB	05 Jun 2025 17:51:50	U	SMPL	
Q2207-37 - 1	BU-703-COMP-05	2486	0.4027	-7471B	PPB	05 Jun 2025 17:54:09	U	SMPL	
Q2208-01 - 1	BU-703-COMP-06	2048	0.3260	-7471B	PPB	05 Jun 2025 17:56:25	U	SMPL	
Q2208-01DUP - 1	BU-703-COMP-06DUP	2081	0.3318	-7471B	PPB	05 Jun 2025 18:01:40	U	SMPL	
Q2208-01MS - 1	BU-703-COMP-06MS	22323	3.8775	-7471B	PPB	05 Jun 2025 18:03:55	U	SMPL	
Q2208-01MSD - 1	BU-703-COMP-06MSD	22398	3.8906	-7471B	PPB	05 Jun 2025 18:06:11	U	SMPL	
Q2208-10 - 1	BU-703-COMP-07	1100	0.1600	-7471B	PPB	05 Jun 2025 18:08:29	U	SMPL	
Q2208-19 - 1	BU-703-COMP-08	1805	0.2835	-7471B	PPB	05 Jun 2025 18:10:48	U	SMPL	
Q2208-28 - 1	BU-703-COMP-09	2996	0.4921	-7471B	PPB	05 Jun 2025 18:13:05	U	SMPL	
Q2214-01 - 1	RBR251425	6157	1.0458	-7471B	PPB	05 Jun 2025 18:15:22	U	SMPL	
Q2216-01 - 1	3898	1135	0.1661	-7471B	PPB	05 Jun 2025 18:17:39	U	SMPL	
CCV06 - 1	CCV06	29226	5.0867	-7471B	PPB	05 Jun 2025 18:19:56	U	SMPL	
CCB06 - 1	CCB06	-91	-0.0487	-7471B	PPB	05 Jun 2025 18:22:11	U	SMPL	
Q2218-01 - 1	72-11934	4356	0.7303	-7471B	PPB	05 Jun 2025 18:24:30	U	SMPL	
Q2219-01 - 1	72-11995	3394	0.5618	-7471B	PPB	05 Jun 2025 18:26:46	U	SMPL	
Q2226-01 - 1	TP06-MHI-WC	2681	0.4369	-7471B	PPB	05 Jun 2025 18:29:02	U	SMPL	
Q2227-01 - 1	TP07-MHI-WC	1195	0.1766	-7471B	PPB	05 Jun 2025 18:31:19	U	SMPL	
Q2228-01 - 1	TP08-MHI-WC	2343	0.3777	-7471B	PPB	05 Jun 2025 18:33:35	U	SMPL	
Q2224-01 - 1		5056	0.8529	-7471B	PPB	05 Jun 2025 18:35:52	U	SMPL	
Q2208-01LX5 - 1		-74	-0.0457	-7471B	PPB	05 Jun 2025 18:40:26	U	SMPL	
Q2208-01A - 1		22132	3.8440	-7471B	PPB	05 Jun 2025 18:42:49	U	SMPL	
HIGH STD - 1	HIGH STD	52138	9.1000	-7471B	PPB	05 Jun 2025 18:46:17	U	SMPL	
CV07 - 1	CCV07	28166	4.9010	-7471B	PPB	05 Jun 2025 18:48:59	U	SMPL	
CB07 - 1	CCB07	9	-0.0311	-7471B	PPB	05 Jun 2025 18:53:54	U	SMPL	