

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136048

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/06/2025	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	06/06/2025	15:11
3	CAL3	1	Water	NA	NA	20.3	10.02	06/06/2025	15:14
4	ICV	1	Water	NA	NA	20.2	7.00	06/06/2025	15:15
5	CCV1	1	Water	NA	NA	20.2	2.01	06/06/2025	15:20
6	Q2246-01	1	Solid	20.02	20	26.6	8.32	06/06/2025	15:25
7	Q2262-01	1	Solid	20.03	20	24.1	10.29	06/06/2025	15:55
8	Q2262-01DUP	1	Solid	20.04	20	24.2	10.30	06/06/2025	15:56
9	Q2262-03	1	Solid	20.02	20	24.1	10.76	06/06/2025	16:00
10	CCV2	1	Water	NA	NA	20.3	12.02	06/06/2025	16:35

WORKLIST(Hardcopy Internal Chain)

US 136048

WorkList Name : PH Q2262

WorkList ID : 190008

Department : Wet-Chemistry

Date : 06-06-2025 14:35:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2246-01	BU-03-060525	Solid	pH	Cool 4 deg C	PSEG05	N51	06/05/2025	9045D
Q2262-01	ARS20-0032	Solid	pH	Cool 4 deg C	PSEG03	N31	06/06/2025	9045D
Q2262-03	ARS20-0001	Solid	pH	Cool 4 deg C	PSEG03	N31	06/06/2025	9045D

Date/Time 06/06/25 15:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06/06/25 18:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]