

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136068

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	06/09/2025	09:15
2	Q2221-02		1	12	100.00	>212.0	>212.0	06/09/2025	09:45
3	Q2221-02DUP		1	13	100.00	>212.0	>212.0	06/09/2025	10:15
4	Q2221-03		1	13	100.00	>212.0	>212.0	06/09/2025	10:45
5	Q2245-01		1	13	100.00	>212.0	>212.0	06/09/2025	11:15
6	Q2245-02		1	12	100.00	>212.0	>212.0	06/09/2025	11:45
7	Q2267-01		1	14	100.00	>212.0	>212.0	06/09/2025	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB136068

WorkList Name : fp-6-9

WorkList ID : 190026

Department : Wet-Chemistry

Date : 06-09-2025 08:45:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2221-02	34900	Water	Flash Point	Cool 4 deg C	PSEG03	N41	06/04/2025	1010B
Q2221-03	34867	Water	Flash Point	Cool 4 deg C	PSEG03	N41	06/04/2025	1010B
Q2245-01	AU-25-0090	Water	Flash Point	Cool 4 deg C	PSEG03	D11	06/05/2025	1010B
Q2245-02	AU-25-0091	Water	Flash Point	Cool 4 deg C	PSEG03	D11	06/05/2025	1010B
Q2267-01	WC-20250605	Water	Flash Point	Cool 4 deg C	PARS02	D41	06/06/2025	1010B

Date/Time 06/09/25 9:05
Raw Sample Received by: 12(wc)
Raw Sample Relinquished by: [Signature]

Date/Time 06/09/25 13:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: 12(wc)