

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136080

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/10/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	06/10/2025	08:02
3	CAL3	1	Water	NA	NA	20.2	10.02	06/10/2025	08:03
4	ICV	1	Water	NA	NA	20.2	7.01	06/10/2025	08:05
5	CCV1	1	Water	NA	NA	20.3	2.01	06/10/2025	08:11
6	Q2271-04	1	Solid	20.02	20	20.7	8.02	06/10/2025	08:20
7	Q2272-04	1	Solid	20.03	20	20.3	7.80	06/10/2025	08:26
8	Q2273-04	1	Solid	20.02	20	20.2	9.20	06/10/2025	08:37
9	Q2273-08	1	Solid	20.01	20	20.5	9.14	06/10/2025	08:45
10	Q2274-04	1	Solid	20.03	20	20.3	5.12	06/10/2025	08:50
11	Q2274-04DUP	1	Solid	20.04	20	20.4	5.13	06/10/2025	08:51
12	CCV2	1	Water	NA	NA	20.2	12.02	06/10/2025	09:00

WORKLIST(Hardcopy Internal Chain)

JB 136080

WorkList Name : corrsivity q2274 WorkList ID : 190053 Department : Wet-Chemistry Date : 06-10-2025 07:34:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2271-04	TP12-MHK-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	06/07/2025	9045D
Q2272-04	TP-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	06/09/2025	9045D
Q2273-04	WC-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	06/09/2025	9045D
Q2273-08	WC-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	06/09/2025	9045D
Q2274-04	TP-13-MHP-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	06/09/2025	9045D

Date/Time 06-10-25 07:50

Raw Sample Received by: SR WCL

Raw Sample Relinquished by: SR WCL

Date/Time

06-10-25

Raw Sample Received by:

SR WCL

Raw Sample Relinquished by:

SR WCL