

16136088

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

6/10/2025 14:21

Reviewed by : RM

Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.933	0.0	0.056	
ICB1	-0.004	0.0	0.006	
CCV1	1.002	0.0	0.059	
CCB1	-0.002	0.0	0.006	
PB168387BL	-0.004	0.0	0.006	
PB168387BS	1.008	0.0	0.060	
Q2126-12	0.022	0.0	0.008	
CCV2	0.998	0.0	0.059	
CCB2	-0.007	0.0	0.006	
N	9			
Mean	0.438			
SD	0.5194			
CV%	118.45			

Aquakem v. 7.2AQ1

Results from time period:

Tue Jun 10 14:11:57 2025

Tue Jun 10 14:17:52 2025

Sample Id	Sam/Ctr/c	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0072	mg/l	6/10/2025 12:36:09	
0.05PPM	A	Phenolics-	P	0.0573	mg/l	6/10/2025 12:36:10	
0.1PPM	A	Phenolics-	P	0.1053	mg/l	6/10/2025 12:36:11	
0.25PPM	A	Phenolics-	P	0.2503	mg/l	6/10/2025 12:36:12	
0.50PPM	A	Phenolics-	P	0.5049	mg/l	6/10/2025 12:36:13	
1.0PPM	A	Phenolics-	P	0.9822	mg/l	6/10/2025 12:36:14	
2.0PPM	A	Phenolics-	P	2.0072	mg/l	6/10/2025 12:36:15	
ICV1	S	Phenolics-	P	0.9329	mg/l	6/10/2025 14:11:58	
ICB1	S	Phenolics-	P	-0.004	mg/l	6/10/2025 14:12:00	
CCV1	S	Phenolics-	P	1.0022	mg/l	6/10/2025 14:12:01	
CCB1	S	Phenolics-	P	-0.0021	mg/l	6/10/2025 14:12:03	
PB168387BL	S	Phenolics-	P	-0.0037	mg/l	6/10/2025 14:12:05	
PB168387BS	S	Phenolics-	P	1.008	mg/l	6/10/2025 14:12:08	
Q2126-12	S	Phenolics-	P	0.022	mg/l	6/10/2025 14:17:48	
CCV2	S	Phenolics-	P	0.9984	mg/l	6/10/2025 14:17:49	
CCB2	S	Phenolics-	P	-0.0074	mg/l	6/10/2025 14:17:51	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

6/10/2025 12:39

Reviewed by : RM

Instrument ID : Konelab


Test Phenolics-

Accepted 6/10/2025 12:39

Factor 18.97
Bias 0.007

Coeff. of det. 0.999834

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.006	-0.0072	0.0000	-
2	Phenol-2	0.010	0.0573	0.0500	14.6
3	Phenol-2	0.012	0.1053	0.1000	5.3
4	Phenol-2	0.020	0.2503	0.2500	0.1
5	Phenol-2	0.033	0.5049	0.5000	1.0
6	Phenol-2	0.058	0.9822	1.0000	-1.0
7	Phenol-2	0.112	2.0072	2.0000	0.4

06/10/2025
RM