

Analytical Summary Report

Analysis Method: SM2510 B
 Parameter: Conductance
 Run Number: LB136091

ANALYST: Eman
 SUPERVISOR REVIEW BY: Iwona
 Meter ID: WC CONDUCTANCE METER-1
 Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP111962

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB136091BL		1	19.6	0.69	0.77	06/10/2025	15:00
2	LB136091BS	1413	1	19.5	1256.00	1403.43	06/10/2025	15:03
3	Q2276-01		1	20.1	878.00	968.66	06/10/2025	15:06
4	Q2276-01DUP		1	20.2	885.00	974.33	06/10/2025	15:09

Note : Conductivity result is corrected to 25°C

Result = (Reading)
 1 + 0.0191 (Temperature - 25)

WORKLIST(Hardcopy Internal Chain)

16136091

WorkList Name : COND-061025 WorkList ID : 190080 Department : Wet-Chemistry Date : 06-10-2025 14:37:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2276-01	MH-6-10-2025	Water	Conductance	Cool 4 deg C	EURO03	D31	06/10/2025	SM2510 B

Date/Time 06/10/25 14:50
Raw Sample Received by: EM (WTC)
Raw Sample Relinquished by: JH (WTC)

Date/Time 06/10/25 15:20
Raw Sample Received by: JH (WTC)
Raw Sample Relinquished by: EM (WTC)