

Analysis Method: 9034
Parameter: Sulfide
Run Number: LB136094

ANALYST: rubina
SUPERVISOR REVIEW BY: Iwona
Constant: 16000
Normality1: 0.025
Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3213
Starch Solution, 4L	W3149

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB168393BL		1	50	50	5.00	0.00	4.92	4.92	0.08	0.00	0.00	06/10/2025	14:45
2	PB168393BS	25	1	50	50	5.00	0.00	1.94	1.94	3.06	2.98	23.84	06/10/2025	14:48
3	Q2126-07		1	50	50	5.00	0.00	4.86	4.86	0.14	0.06	0.48	06/10/2025	14:51
4	Q2126-08		1	50	50	5.00	0.00	4.78	4.78	0.22	0.14	1.12	06/10/2025	14:54

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume