

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136095

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/11/2025	08:35
2	Q2221-01		1	5	7.00	44.6	44.3	06/11/2025	09:05
3	Q2221-01DUP		1	5	7.00	44.6	44.3	06/11/2025	09:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136095

WorkList Name : FP-6-10

WorkList ID : 190083

Department : Wet-Chemistry

Date : 06-10-2025 13:04:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2221-01	34798	Solid	Flash Point	Cool 4 deg C	PSEG03	N41	06/04/2025	1010B

Date/Time 26/11/2025 08.25
 Raw Sample Received by: RM(wc)
 Raw Sample Relinquished by: JH Coo

Date/Time 26/11/2025 10.15
 Raw Sample Received by: JH Coo
 Raw Sample Relinquished by: RM(wc)