

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136098

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.4	06/10/2025	15:00
2	Q2279-01		1	14	100.00	>212.0	>212.0	06/10/2025	15:30
3	Q2279-01DUP		1	14	100.00	>212.0	>212.0	06/10/2025	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB136098

WorkList Name : fp-6-10-

WorkList ID : 190094

Department : Wet-Chemistry

Date : 06-10-2025 14:05:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2279-01	301-469-5TH-AVE	Water	Flash Point	Cool 4 deg C	PSEG03	D41	06/10/2025	1010B

Date/Time 06/10/2025 14.40
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: M cwc

Date/Time 06/10/2025 16.40
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: RM cwc