

LB136099


245.1

INSTRUMENT ID: CXL

Linear

 μ Abs.:

39755



A= 0.0000e+000

B= 2.5282e-004

C= 1.8202e-002

Rho= 0.9999259

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.0	0.000	-0.011	-0.011	-115	0.000	-115					
0.05	0.050	0.048	-0.002	116	0.0 %	116	0				-4
0.2	0.200	0.241	0.041	881	0.0 %	881					21
2.5	2.500	2.522	0.022	9905	0.0 %	9905					1
5.0	5.000	4.950	-0.050	19509	0.0 %	19509					1
7.5	7.500	7.431	-0.069	29320	0.0 %	29320					1
10.0	10.000	10.069	0.069	39755	0.0 %	39755					1

LB136099 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Jun 2025 08:59:48

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	-115	-	0.0000	245.1	PPB	11 Jun 2025 09:13:30	S	Std
0.05 - 1	50.45	116	-	0.0500	245.1	PPB	11 Jun 2025 09:15:48	S	Std
0.2 - 1	50.2	881	-	0.2000	245.1	PPB	11 Jun 2025 09:18:06	S	Std
2.5 - 1	52.5	9905	-	2.5000	245.1	PPB	11 Jun 2025 09:20:23	S	Std
5.0 - 1	55	19509	-	5.0000	245.1	PPB	11 Jun 2025 09:22:40	S	Std
7.5 - 1	57.5	29320	-	7.5000	245.1	PPB	11 Jun 2025 09:24:57	S	Std
10.0 - 1	510	39755	-	10.0000	245.1	PPB	11 Jun 2025 09:40:38	S	Std
ICV32 - 1	ICV32	16378	4.1588	-245.1		PPB	11 Jun 2025 09:46:06	U	SMPL
ICB32 - 1	ICB32	-328	-0.0647	-245.1		PPB	11 Jun 2025 09:48:21	U	SMPL
CCV23 - 1	CCV23	20002	5.0750	-245.1		PPB	11 Jun 2025 09:50:38	U	SMPL
CCB23 - 1	CCB23	-419	-0.0877	-245.1		PPB	11 Jun 2025 09:52:54	U	SMPL
CRA - 1	CRA	99	0.0432	-245.1		PPB	11 Jun 2025 09:55:11	U	SMPL
HighStd - 1	HighStd	39762	10.0707	-245.1		PPB	11 Jun 2025 09:57:26	U	SMPL
ChkStd - 1	ChkStd	25321	6.4198	-245.1		PPB	11 Jun 2025 09:59:41	U	SMPL
PB168402BL - 1	PBW	-53	0.0048	-245.1		PPB	11 Jun 2025 10:04:48	U	SMPL
PB168402BS - 1	LCSW	15515	3.9406	-245.1		PPB	11 Jun 2025 10:07:04	U	SMPL
Q2148-02 - 1	COMP	407	0.1211	-245.1		PPB	11 Jun 2025 10:09:20	U	SMPL
Q2264-04 - 1	EF-WW	83	0.0392	-245.1		PPB	11 Jun 2025 10:11:39	U	SMPL
Q2264-04DUP - 1	EF-WWDUP	85	0.0397	-245.1		PPB	11 Jun 2025 10:13:56	U	SMPL
Q2264-04MS - 1	EF-WWMS	14210	3.6107	-245.1		PPB	11 Jun 2025 10:18:51	U	SMPL
Q2264-04MSD - 1	EF-WWMSD	14330	3.6411	-245.1		PPB	11 Jun 2025 10:21:07	U	SMPL
CCV24 - 1	CCV24	19735	5.0075	-245.1		PPB	11 Jun 2025 10:23:25	U	SMPL
CCB24 - 1	CCB24	-16	0.0142	-245.1		PPB	11 Jun 2025 10:28:24	U	SMPL
Q2264-04LX5 - 1		-11	0.0154	-245.1		PPB	11 Jun 2025 10:30:38	U	SMPL
Q2264-04A - 1		15091	3.8334	-245.1		PPB	11 Jun 2025 10:32:54	U	SMPL
CCV25 - 1	CCV25	19384	4.9188	-245.1		PPB	11 Jun 2025 10:35:10	U	SMPL
CCB25 - 1	CCB25	-276	-0.0516	-245.1		PPB	11 Jun 2025 10:37:28	U	SMPL