

## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136102

Slope : 98.4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W3178         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W3093         |
| PH 10.01 BUFFER,COLOR CD 475ML | W3191         |
| buffer solution pH 7 yellow    | W3071         |
| Buffer Solution, PH2 (500ml)   | W3161         |
| pH 12.00 Buffer                | W3200         |

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 06/11/2025 | 11:30     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 06/11/2025 | 11:31     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.2             | 10.02       | 06/11/2025 | 11:33     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 06/11/2025 | 11:35     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 06/11/2025 | 11:36     |
| 6   | Q2276-01    | 1  | Water  | NA          | NA          | 20.1             | 3.76        | 06/11/2025 | 11:45     |
| 7   | Q2279-01    | 1  | Water  | NA          | NA          | 20.1             | 7.77        | 06/11/2025 | 12:00     |
| 8   | Q2282-03    | 1  | Water  | NA          | NA          | 20.5             | 5.30        | 06/11/2025 | 12:12     |
| 9   | Q2282-04    | 1  | Water  | NA          | NA          | 20.4             | 7.35        | 06/11/2025 | 12:20     |
| 10  | Q2284-02    | 1  | Water  | NA          | NA          | 20.1             | 9.28        | 06/11/2025 | 12:30     |
| 11  | Q2286-03    | 1  | Water  | NA          | NA          | 20.1             | 6.05        | 06/11/2025 | 12:37     |
| 12  | Q2286-03DUP | 1  | Water  | NA          | NA          | 20.2             | 6.06        | 06/11/2025 | 12:40     |
| 13  | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.02       | 06/11/2025 | 12:41     |