

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136123

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/12/2025	08:10
2	CAL2	1	Water	NA	NA	20.1	7.01	06/12/2025	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	06/12/2025	08:14
4	ICV	1	Water	NA	NA	20.2	7.01	06/12/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	06/12/2025	08:20
6	Q2288-01	1	Solid	20.02	20	20.7	7.21	06/12/2025	08:30
7	Q2288-01DUP	1	Solid	20.03	20	20.9	7.22	06/12/2025	08:33
8	CCV2	1	Water	NA	NA	20.2	12.02	06/12/2025	08:35

WORKLIST(Hardcopy Internal Chain)

JB 136123

WorkList Name : PH W Q2289

WorkList ID : 190119

Department : Wet-Chemistry

Date : 06-12-2025 07:58:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2288-01	ETGI-329	Solid	pH	Cool 4 deg C	PSEG03	D31	06/10/2025	9045D

Date/Time 06-12-25 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06-12-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]