

LB136124


7470A

INSTRUMENT ID! EVZ

Linear

μ Abs.:

42381



A= 0.0000e+000

B= 2.3760e-004

C= -1.6936e-002

Rho= 0.9997880

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/11
0.0	0.000	0.002	0.002	80	0.000	80	0				12
0.2	0.200	0.224	0.024	1016	0.0 %	1016					3
2.5	2.500	2.575	0.075	10907	0.0 %	10907					2
5.0	5.000	4.840	-0.160	20441	0.0 %	20441					0
7.5	7.500	7.506	0.006	31662	0.0 %	31662					1
10.0	10.000	10.053	0.053	42381	0.0 %	42381					

LB136124 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jun 2025 09:14:22

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	80	-	0.0000	7470A	PPB	12 Jun 2025 09:52:49	S	Std
0.2 - 1	50.2	1016	-	0.2000	7470A	PPB	12 Jun 2025 09:55:06	S	Std
2.5 - 1	52.5	10907	-	2.5000	7470A	PPB	12 Jun 2025 09:57:23	S	Std
5.0 - 1	55	20441	-	5.0000	7470A	PPB	12 Jun 2025 09:59:40	S	Std
7.5 - 1	57.5	31662	-	7.5000	7470A	PPB	12 Jun 2025 10:01:58	S	Std
10.0 - 1	59.0	42381	-	10.0000	7470A	PPB	12 Jun 2025 10:04:16	S	Std
ICV35 - 1	ICV35	16958	4.0123	-	7470A	PPB	12 Jun 2025 10:12:22	U	SMPL
ICB35 - 1	ICB35	-249	-0.0761	-	7470A	PPB	12 Jun 2025 10:14:38	U	SMPL
CCV30 - 1	CCV30	19587	4.6370	-	7470A	PPB	12 Jun 2025 10:16:55	U	SMPL
CCB30 - 1	CCB30	-294	-0.0868	-	7470A	PPB	12 Jun 2025 10:19:10	U	SMPL
CRA - 1	CRA	852	0.1855	-	7470A	PPB	12 Jun 2025 10:21:27	U	SMPL
HighStd - 1	HighStd	40841	9.6870	-	7470A	PPB	12 Jun 2025 10:23:42	U	SMPL
ChkStd - 1	ChkStd	26316	6.2358	-	7470A	PPB	12 Jun 2025 10:25:57	U	SMPL
PB168424BL - 1	PBW	-336	-0.0968	-	7470A	PPB	12 Jun 2025 10:28:16	U	SMPL
PB168424BS - 1	LCSW	15010	3.5495	-	7470A	PPB	12 Jun 2025 10:30:35	U	SMPL
Q2259-02 - 1	OU4-PCS-TC-36-060525	1016	0.2245	-	7470A	PPB	12 Jun 2025 10:32:52	U	SMPL
Q2259-04 - 1	OU4-PCS-TC-37-060525	-15	-0.0205	-	7470A	PPB	12 Jun 2025 10:35:10	U	SMPL
Q2259-04DUP - 1	OU4-PCS-TC-37-060525DUP	166	0.0225	-	7470A	PPB	12 Jun 2025 10:37:29	U	SMPL
Q2259-04MS - 1	OU4-PCS-TC-37-060525MS	14410	3.4069	-	7470A	PPB	12 Jun 2025 10:39:46	U	SMPL
Q2259-04MSD - 1	OU4-PCS-TC-37-060525MSD	13557	3.2042	-	7470A	PPB	12 Jun 2025 10:42:07	U	SMPL
CCV31 - 1	CCV31	21296	5.0431	-	7470A	PPB	12 Jun 2025 10:50:22	U	SMPL
CCB31 - 1	CCB31	-306	-0.0896	-	7470A	PPB	12 Jun 2025 10:52:37	U	SMPL
PB168425BL - 1	PBW	-30	-0.0241	-	7470A	PPB	12 Jun 2025 10:54:58	U	SMPL
PB168425BS - 1	LCSW	15541	3.6757	-	7470A	PPB	12 Jun 2025 10:57:10	U	SMPL
Q2267-01 - 1	WC-20250605	-144	-0.0512	-	7470A	PPB	12 Jun 2025 10:59:26	U	SMPL
Q2271-04 - 1	TP12-MHK-WC	-17	-0.0210	-	7470A	PPB	12 Jun 2025 11:01:45	U	SMPL
Q2272-04 - 1	TP-6	16	-0.0131	-	7470A	PPB	12 Jun 2025 11:04:02	U	SMPL
Q2273-04 - 1	WC-4	-12	-0.0198	-	7470A	PPB	12 Jun 2025 11:06:19	U	SMPL
Q2273-08 - 1	WC-6	-9	-0.0191	-	7470A	PPB	12 Jun 2025 11:08:35	U	SMPL
Q2274-04 - 1	TP-13-MHP-WC	1	-0.0167	-	7470A	PPB	12 Jun 2025 11:10:52	U	SMPL
Q2278-04 - 1	TP-2	15	-0.0134	-	7470A	PPB	12 Jun 2025 11:13:14	U	SMPL
Q2280-02 - 1	VNJ-210	14	-0.0136	-	7470A	PPB	12 Jun 2025 11:15:35	U	SMPL
CCV32 - 1	CCV32	19895	4.7102	-	7470A	PPB	12 Jun 2025 11:17:50	U	SMPL
CCB32 - 1	CCB32	-248	-0.0759	-	7470A	PPB	12 Jun 2025 11:20:06	U	SMPL
Q2280-04 - 1	RT-4643	6	-0.0155	-	7470A	PPB	12 Jun 2025 11:22:24	U	SMPL
Q2285-05 - 1	HAM-CONCRETE	-2	-0.0174	-	7470A	PPB	12 Jun 2025 11:24:39	U	SMPL
Q2285-05DUP - 1	HAM-CONCRETEDUP	0	-0.0169	-	7470A	PPB	12 Jun 2025 11:26:55	U	SMPL
Q2285-05MS - 1	HAM-CONCRETEMS	12543	2.9633	-	7470A	PPB	12 Jun 2025 11:29:10	U	SMPL
Q2285-05MSD - 1	HAM-CONCRETEMSD	12080	2.8533	-	7470A	PPB	12 Jun 2025 11:31:26	U	SMPL
CCV33 - 1	CCV33	20231	4.7900	-	7470A	PPB	12 Jun 2025 11:38:21	U	SMPL
CCB33 - 1	CCB33	-252	-0.0768	-	7470A	PPB	12 Jun 2025 11:40:37	U	SMPL
PB168362TB - 1	PB168362TB	5	-0.0157	-	7470A	PPB	12 Jun 2025 11:42:57	U	SMPL
PB168369TB - 1	PB168369TB	1	-0.0167	-	7470A	PPB	12 Jun 2025 11:45:14	U	SMPL
PB168390TB - 1	PB168390TB	6	-0.0155	-	7470A	PPB	12 Jun 2025 11:47:29	U	SMPL
Q2259-04LX5 - 1		300	0.0543	-	7470A	PPB	12 Jun 2025 11:49:45	U	SMPL
Q2259-04A - 1		13340	3.1527	-	7470A	PPB	12 Jun 2025 11:52:00	U	SMPL
Q2285-05LX5 - 1		-217	-0.0685	-	7470A	PPB	12 Jun 2025 11:54:17	U	SMPL
Q2285-05A - 1		13696	3.2373	-	7470A	PPB	12 Jun 2025 11:56:35	U	SMPL
CCV34 - 1	CCV34	20632	4.8853	-	7470A	PPB	12 Jun 2025 11:58:51	U	SMPL
CCB34 - 1	CCB34	-321	-0.0932	-	7470A	PPB	12 Jun 2025 12:01:09	U	SMPL