



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/13/2025

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 06/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 06/13/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136126

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2287-01	CONCRETE-SLAB	1	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
Q2296-01	WC-1	2	1.18	10.56	11.74	9.74	81.1	
Q2296-02	WC-1-EPH	3	1.18	10.23	11.41	9.55	81.8	
Q2296-03	WC-1-VOC	4	1.19	10.57	11.76	9.63	79.8	
Q2296-05	WC-2	5	1.19	10.58	11.77	10.52	88.2	
Q2296-06	WC-2-EPH	6	1.19	10.30	11.49	10.5	90.4	
Q2296-07	WC-2-VOC	7	1.19	10.21	11.4	10.62	92.4	
Q2296-09	WC-3	8	1.15	10.78	11.93	10.33	85.2	
Q2296-10	WC-3-EPH	9	1.19	10.60	11.79	10.1	84.1	
Q2296-11	WC-3-VOC	10	1.18	10.81	11.99	10.36	84.9	
Q2296-13	WC-4	11	1.15	10.77	11.92	9.95	81.7	
Q2296-14	WC-4-EPH	12	1.13	10.48	11.61	9.88	83.5	
Q2296-15	WC-4-VOC	13	1.14	10.72	11.86	9.71	79.9	
Q2296-17	WC-5	14	1.17	10.49	11.66	10.63	90.2	
Q2296-18	WC-5-EPH	15	1.17	10.50	11.67	10.32	87.1	
Q2296-19	WC-5-VOC	16	1.15	10.83	11.98	10.97	90.7	
Q2296-21	WC-6	19	1.13	10.37	11.5	10.2	87.5	
Q2296-22	WC-6-EPH	17	1.14	10.44	11.58	10.65	91.1	
Q2296-23	WC-6-VOC	18	1.17	10.82	11.99	10.36	84.9	
Q2297-01	TP-3	20	1.15	10.96	12.11	10.2	82.6	
Q2297-02	TP-3-EPH	21	1.14	10.85	11.99	11.04	91.2	
Q2297-03	TP-3-VOC	22	1.12	10.16	11.28	10.3	90.4	
Q2298-01	AU-05-061125	23	1.15	11.63	12.78	11.7	90.7	
Q2298-02	AU-05-061125	24	1.15	11.63	12.78	11.7	90.7	
Q2301-02	WC-URBAN-FILL-C	25	1.16	10.37	11.53	9.77	83.0	
Q2303-01	B-165-SB01	26	1.14	10.74	11.88	10.12	83.6	
Q2303-02	B-170-SB03	27	1.13	10.37	11.5	10.74	92.7	
Q2304-01	RBR200057-1	28	1.00	1.00	2.00	2.00	100.0	wipe sample



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/13/2025

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 06/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 06/13/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136126

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2304-02	RBR200057-2	29	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2304-03	VNJ239-3	30	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2304-04	VNJ239-4	31	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2304-05	VNJ239-5	32	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2305-01	TR-04-06122025	33	1.18	10.17	11.35	11.17	98.2	
Q2305-02	TR-04-06122025-E2	34	1.13	10.69	11.82	11.24	94.6	
Q2307-01	LINDEN-SAA	35	1.14	10.54	11.68	8.87	73.3	
Q2307-02	LINDEN-SAA	36	1.13	10.70	11.83	8.87	72.3	
Q2308-01	EO-02-06122025	37	1.12	10.24	11.36	10.38	90.4	
Q2308-02	EO-02-06122025-E2	38	1.19	10.75	11.94	10.98	91.1	
Q2310-01	TP-7	39	1.18	10.42	11.6	10.42	88.7	
Q2310-02	TP-7-EFH	40	1.15	11.29	12.44	11.08	88.0	
Q2310-03	TP-7-VOC	41	1.12	10.91	12.03	10.99	90.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$