

LB136130


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

49444



A= 0.0000e+000

B= 2.0058e-004

C= -4.4244e-003

Rho= 0.9996714

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.007	0.007	57	0.000	57					0.00
0.2	0.200	0.223	0.023	1133	0.0 %	1133					12
2.5	2.500	2.475	-0.025	12360	0.0 %	12360					-1
5.0	5.000	4.897	-0.103	24438	0.0 %	24438					-2
7.5	7.500	7.685	0.185	38338	0.0 %	38338					2
10.0	10.000	9.913	-0.087	49444	0.0 %	49444					-1

LB136130 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Jun 2025 13:46:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7471B	PPB	12 Jun 2025 13:47:06	S	Std
0.2 - 1	502	1133	-	0.2000	7471B	PPB	12 Jun 2025 13:49:23	S	Std
2.5 - 1	52.5	12360	-	2.5000	7471B	PPB	12 Jun 2025 13:51:40	S	Std
5.0 - 1	55	24438	-	5.0000	7471B	PPB	12 Jun 2025 13:53:57	S	Std
7.5 - 1	59.5	38338	-	7.5000	7471B	PPB	12 Jun 2025 13:56:16	S	Std
10.0 - 1	510	49444	-	10.0000	7471B	PPB	12 Jun 2025 14:01:56	S	Std
ICV36 - 1	ICV36	20328	4.0729	-	7471B	PPB	12 Jun 2025 14:04:54	U	SMPL
ICB36 - 1	ICB36	-307	-0.0660	-	7471B	PPB	12 Jun 2025 14:07:10	U	SMPL
CCV35 - 1	CCV35	24539	4.9175	-	7471B	PPB	12 Jun 2025 14:09:28	U	SMPL
CCB35 - 1	CCB35	-289	-0.0624	-	7471B	PPB	12 Jun 2025 14:11:44	U	SMPL
CRA - 1	CRA	1084	0.2130	-	7471B	PPB	12 Jun 2025 14:14:04	U	SMPL
HighStd - 1	HighStd	49096	9.8431	-	7471B	PPB	12 Jun 2025 14:16:19	U	SMPL
ChkStd - 1	ChkStd	35080	7.0318	-	7471B	PPB	12 Jun 2025 14:18:35	U	SMPL
PB168449BL - 1	PBS	-393	-0.0833	-	7471B	PPB	12 Jun 2025 14:21:12	U	SMPL
PB168449BS - 1	LCSS	19863	3.9796	-	7471B	PPB	12 Jun 2025 14:26:57	U	SMPL
Q2271-01 - 1	TP12-MHK-WC	2639	0.5249	-	7471B	PPB	12 Jun 2025 14:29:13	U	SMPL
Q2272-01 - 1	TP-6	814	0.1588	-	7471B	PPB	12 Jun 2025 14:31:34	U	SMPL
Q2273-01 - 1	WC-4	79	0.0114	-	7471B	PPB	12 Jun 2025 14:33:53	U	SMPL
Q2273-05 - 1	WC-6	94	0.0144	-	7471B	PPB	12 Jun 2025 14:36:09	U	SMPL
Q2274-01 - 1	TP-13-MHP-WC	2461	0.4892	-	7471B	PPB	12 Jun 2025 14:38:24	U	SMPL
CCV36 - 1	CCV36	23169	4.6427	-	7471B	PPB	12 Jun 2025 14:55:12	U	SMPL
CCB36 - 1	CCB36	-326	-0.0698	-	7471B	PPB	12 Jun 2025 14:57:29	U	SMPL
Q2278-01 - 1	TP-2	11978	2.3981	-	7471B	PPB	12 Jun 2025 14:59:50	U	SMPL
Q2280-01 - 1	VNJ-210	2047	0.4062	-	7471B	PPB	12 Jun 2025 15:02:05	U	SMPL
Q2280-03 - 1	RT-4643	3036	0.6045	-	7471B	PPB	12 Jun 2025 15:04:28	U	SMPL
Q2280-03DUP - 1	RT-4643DUP	3265	0.6505	-	7471B	PPB	12 Jun 2025 15:06:45	U	SMPL
Q2280-03MS - 1	RT-4643MS	24827	4.9753	-	7471B	PPB	12 Jun 2025 15:13:48	U	SMPL
Q2280-03MSD - 1	RT-4643MSD	24941	4.9982	-	7471B	PPB	12 Jun 2025 15:16:06	U	SMPL
Q2280-0LX53 - 1		207	0.0371	-	7471B	PPB	12 Jun 2025 15:18:25	U	SMPL
Q2280-03A - 1		25395	5.0892	-	7471B	PPB	12 Jun 2025 15:20:45	U	SMPL
CCV37 - 1	CCV37	26919	5.3949	-	7471B	PPB	12 Jun 2025 15:26:21	U	SMPL
CCB37 - 1	CCB37	-85	-0.0215	-	7471B	PPB	12 Jun 2025 15:28:36	U	SMPL