

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136141

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/13/2025	08:05
2	CAL2	1	Water	NA	NA	20.2	7.00	06/13/2025	08:06
3	CAL3	1	Water	NA	NA	20.2	10.02	06/13/2025	08:09
4	ICV	1	Water	NA	NA	20.1	7.01	06/13/2025	08:10
5	CCV1	1	Water	NA	NA	20.2	2.02	06/13/2025	08:14
6	Q2307-03	1	Water	NA	NA	23.6	8.01	06/13/2025	08:30
7	Q2307-03DUP	1	Water	NA	NA	23.8	8.02	06/13/2025	08:34
8	CCV2	1	Water	NA	NA	20.2	12.02	06/13/2025	08:35

WORKLIST(Hardcopy Internal Chain)

VB 136147

WorkList Name : PH W Q2307 WorkList ID : 190166 Department : Wet-Chemistry Date : 06-13-2025 07:46:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2307-03	B LINDEN-SAA-WATER	Water	pH	Cool 4 deg C	PSEG03	D51	06/12/2025	9040C

Date/Time 06/13/25 07:57
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06/13/25 10:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]