

LB136146


7471B

(INSTRUMENT ID: CV1

Linear

μ Abs.:

59856



A= 0.0000e+000

B= 1.6800e-004

C= 3.5638e-002

Rho= 0.9994899

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.019	0.019	-100	0.000	-100					0
0.2	0.200	0.230	0.030	1156	0.0 %	1156					15
2.5	2.500	2.614	0.114	15346	0.0 %	15346					5
5.0	5.000	4.754	-0.246	28083	0.0 %	28083					5
7.5	7.500	7.493	-0.007	44387	0.0 %	44387					0
10.0	10.000	10.091	0.091	59856	0.0 %	59856					1

LB1314666666 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jun 2025 10:02:42

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-100	-	0.0007471B	PPB		13 Jun 2025 10:16:07	S	Std
0.2 - 1	50.2	1156	-	0.20007471B	PPB		13 Jun 2025 10:18:24	S	Std
2.5 - 1	50.5	15346	-	2.50007471B	PPB		13 Jun 2025 10:20:41	S	Std
5.0 - 1	50.5	28083	-	5.00007471B	PPB		13 Jun 2025 10:22:58	S	Std
7.5 - 1	50.5	44387	-	7.50007471B	PPB		13 Jun 2025 10:25:18	S	Std
10.0 - 1	50.5	58856	-	10.00007471B	PPB		13 Jun 2025 10:27:38	S	Std
ICV37 - 1	ICV37	21945	3.7224	-7471B	PPB		13 Jun 2025 10:30:41	U	SMPL
ICB37 - 1	ICB37	-213	-0.0001	-7471B	PPB		13 Jun 2025 10:32:57	U	SMPL
CCV38 - 1	CCV38	28255	4.7825	-7471B	PPB		13 Jun 2025 10:35:16	U	SMPL
CCB38 - 1	CCB38	-400	-0.0316	-7471B	PPB		13 Jun 2025 10:37:33	U	SMPL
CRA - 1	CRA	1136	0.2265	-7471B	PPB		13 Jun 2025 10:40:06	U	SMPL
HighStd - 1	HighStd	59725	10.0694	-7471B	PPB		13 Jun 2025 10:42:22	U	SMPL
ChkStd - 1	ChkStd	37733	6.3747	-7471B	PPB		13 Jun 2025 10:44:37	U	SMPL
PB168454BL - 1	PBS	-28	0.0309	-7471B	PPB		13 Jun 2025 10:49:40	U	SMPL
PB168454BS - 1	LCSS	22181	3.7620	-7471B	PPB		13 Jun 2025 10:55:20	U	SMPL
Q2126-01 - 1	LOD-MDL-SOIL-03-QT2-2025	90	0.0508	-7471B	PPB		13 Jun 2025 11:09:54	U	SMPL
Q2126-02 - 1	LOQ-SOIL-02-QT2-2025	1172	0.2325	-7471B	PPB		13 Jun 2025 11:12:11	U	SMPL
Q2283-01 - 1	MOO-25-0166-MOO-25-0167	3877	0.6870	-7471B	PPB		13 Jun 2025 11:14:28	U	SMPL
Q2286-01 - 1	LAW-25-0082-LAW-250083	9424	1.6189	-7471B	PPB		13 Jun 2025 11:16:45	U	SMPL
Q2287-01 - 1	CONCRETE-SLAB	-3	0.0351	-7471B	PPB		13 Jun 2025 11:19:03	U	SMPL
CCV39 - 1	CCV39	28217	4.7761	-7471B	PPB		13 Jun 2025 11:21:21	U	SMPL
CCB39 - 1	CCB39	-358	-0.0245	-7471B	PPB		13 Jun 2025 11:23:36	U	SMPL
Q2295-01 - 1	TP01-MH1-WC	2027	0.3762	-7471B	PPB		13 Jun 2025 11:25:56	U	SMPL
Q2295-05 - 1	TP02-MH5-WC	3849	0.6823	-7471B	PPB		13 Jun 2025 11:28:11	U	SMPL
Q2296-01 - 1	WC-1	5515	0.9622	-7471B	PPB		13 Jun 2025 11:30:28	U	SMPL
Q2296-05 - 1	WC-2	5917	1.0297	-7471B	PPB		13 Jun 2025 11:32:45	U	SMPL
Q2296-09 - 1	WC-3	1032	0.2090	-7471B	PPB		13 Jun 2025 11:35:02	U	SMPL
Q2296-13 - 1	WC-4	2253	0.4141	-7471B	PPB		13 Jun 2025 11:37:21	U	SMPL
Q2296-17 - 1	WC-5	728	0.1579	-7471B	PPB		13 Jun 2025 11:39:38	U	SMPL
Q2296-21 - 1	WC-6	7960	1.3729	-7471B	PPB		13 Jun 2025 11:41:56	U	SMPL
Q2297-01 - 1	TP-3	5538	0.9660	-7471B	PPB		13 Jun 2025 11:56:57	U	SMPL
Q2297-01DUP - 1	TP-3DUP	5304	0.9267	-7471B	PPB		13 Jun 2025 11:59:13	U	SMPL
CCV40 - 1	CCV40	28383	4.8040	-7471B	PPB		13 Jun 2025 12:01:31	U	SMPL
CCB40 - 1	CCB40	-322	-0.0185	-7471B	PPB		13 Jun 2025 12:03:47	U	SMPL
Q2297-01MS - 1	TP-3MS	30849	5.2182	-7471B	PPB		13 Jun 2025 12:06:07	U	SMPL
Q2297-01MSD - 1	TP-3MSD	30077	5.0885	-7471B	PPB		13 Jun 2025 12:13:38	U	SMPL
Q2298-01 - 1	AU-05-061125	2513	0.4578	-7471B	PPB		13 Jun 2025 12:15:53	U	SMPL
Q2305-01 - 1	TR-04-06122025	7956	1.3722	-7471B	PPB		13 Jun 2025 12:18:17	U	SMPL
Q2307-01 - 1	LINDEN-SAA	3180	0.5699	-7471B	PPB		13 Jun 2025 12:20:34	U	SMPL
Q2308-01 - 1	EO-02-06122025	2605	0.4733	-7471B	PPB		13 Jun 2025 12:22:53	U	SMPL
Q2310-01 - 1	TP-7	395	0.1020	-7471B	PPB		13 Jun 2025 12:25:10	U	SMPL
Q2297-01LX5 - 1		985	0.2011	-7471B	PPB		13 Jun 2025 12:27:27	U	SMPL
Q2297-01A - 1		29928	5.0635	-7471B	PPB		13 Jun 2025 12:29:44	U	SMPL
CCV41 - 1	CCV41	28832	4.8794	-7471B	PPB		13 Jun 2025 12:32:00	U	SMPL
CCB41 - 1	CCB41	-393	-0.0304	-7471B	PPB		13 Jun 2025 12:34:21	U	SMPL