



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/17/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 06/16/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 06/17/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136165

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2314-08	BP-VPB-182-GW-840-842	25	1.15	8.70	9.85	1.84	7.9	sludge sample
Q2335-01	ETGI-371-1	1	1.00	1.00	2.00	2.00	100.0	WIPE
Q2335-02	ETGI-371-2	2	1.00	1.00	2.00	2.00	100.0	WIPE
Q2335-03	ETGI-361-1	3	1.00	1.00	2.00	2.00	100.0	WIPE
Q2335-04	ETGI-361-2	4	1.00	1.00	2.00	2.00	100.0	WIPE
Q2336-01	WBR-1	5	1.16	8.61	9.77	7.68	75.7	
Q2336-02	WBR-1-E2	6	1.15	8.45	9.6	7.68	77.3	
Q2336-03	WBR-2	7	1.15	8.79	9.94	8.4	82.5	
Q2336-04	WBR-2-E2	8	1.17	8.57	9.74	8.26	82.7	
Q2337-01	OR-02-061625	9	1.16	8.74	9.9	9.25	92.6	
Q2337-02	OR-02-061625-E2	10	1.16	8.56	9.72	8.88	90.2	
Q2338-01	OK-03-061625	11	1.18	8.45	9.63	8.92	91.6	
Q2338-02	OK-03-061625-E2	12	1.15	8.50	9.65	8.88	90.9	
Q2339-01	MH-C	13	1.16	8.69	9.85	8.96	89.8	
Q2339-02	MH-C-EPH	14	1.15	8.43	9.58	8.7	89.6	
Q2339-03	MH-C-VOC	15	1.16	8.43	9.59	8.71	89.6	
Q2340-01	TP05-MH17A-WC	16	1.17	8.70	9.87	9.06	90.7	
Q2340-02	TP05-MH17A-VOC	17	1.15	8.80	9.95	9.2	91.5	
Q2340-03	TP05-MH17A-EPH	18	1.14	8.68	9.82	9.11	91.8	
Q2341-01	TP-9	19	1.16	8.55	9.71	8.88	90.3	
Q2341-02	TP-9-EPH	20	1.16	8.61	9.77	8.9	89.9	
Q2341-03	TP-9-VOC	21	1.15	96.95	98.1	8.93	8.0	
Q2341-05	EP-3	22	1.14	8.61	9.75	8.98	91.1	
Q2341-06	EP-3-EPH	23	1.15	8.73	9.88	9.00	89.9	
Q2341-07	EP-3-VOC	24	1.14	8.59	9.73	8.91	90.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$