

LB136169


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5361E



A= 0.0000e+000

B= 1.8389e-004

C= -4.9349e-002

Rho= 0.9993808

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.0	0.000	-0.011	-0.011	209	0.000	209					1
0.2	0.200	0.201	0.001	1361	0.0 %	1361					1
2.5	2.500	2.420	-0.080	13427	0.0 %	13427					3
5.0	5.000	5.041	0.041	27682	0.0 %	27682					1
7.5	7.500	7.739	0.239	42350	0.0 %	42350					4
10.0	10.000	9.810	-0.190	53616	0.0 %	53616					2

LB136169 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Jun 2025 10:26:39

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	209	-	0.0000	7471B	PPB	17 Jun 2025 10:42:02	S	Std
0.2 - 1	502	1361	-	0.2000	7471B	PPB	17 Jun 2025 10:44:19	S	Std
2.5 - 1	52.5	13427	-	2.5000	7471B	PPB	17 Jun 2025 10:46:37	S	Std
5.0 - 1	55	27682	-	5.0000	7471B	PPB	17 Jun 2025 10:48:54	S	Std
7.5 - 1	57.5	42350	-	7.5000	7471B	PPB	17 Jun 2025 10:51:13	S	Std
10.0 - 1	57.5	53616	-	10.0000	7471B	PPB	17 Jun 2025 11:05:09	S	Std
ICV42 - 1	ICV42	21520	3.9081	-	7471B	PPB	17 Jun 2025 11:10:09	U	SMPL
ICB42 - 1	ICB42	-208	-0.0876	-	7471B	PPB	17 Jun 2025 11:12:24	U	SMPL
CCV53 - 1	CCV53	27789	5.0609	-	7471B	PPB	17 Jun 2025 11:14:42	U	SMPL
CCB53 - 1	CCB53	-240	-0.0935	-	7471B	PPB	17 Jun 2025 11:16:57	U	SMPL
CRA - 1	CRA	1248	0.1802	-	7471B	PPB	17 Jun 2025 11:19:18	U	SMPL
HighStd - 1	HighStd	54052	9.8905	-	7471B	PPB	17 Jun 2025 11:21:33	U	SMPL
ChkStd - 1	ChkStd	41586	7.5981	-	7471B	PPB	17 Jun 2025 11:26:53	U	SMPL
PB168512BL - 1	PBS	-19	-0.0528	-	7471B	PPB	17 Jun 2025 11:32:34	U	SMPL
PB168512BS - 1	LCSS	21925	3.9825	-	7471B	PPB	17 Jun 2025 11:37:24	U	SMPL
Q2312-01 - 1	TP-1	1138	0.1599	-	7471B	PPB	17 Jun 2025 11:39:40	U	SMPL
Q2312-01DUP - 1	TP-1DUP	1099	0.1528	-	7471B	PPB	17 Jun 2025 11:45:53	U	SMPL
Q2312-01MS - 1	TP-1MS	22222	4.0372	-	7471B	PPB	17 Jun 2025 11:48:10	U	SMPL
Q2312-01MSD - 1	TP-1MSD	21931	3.9836	-	7471B	PPB	17 Jun 2025 11:50:27	U	SMPL
Q2322-01 - 1	CL-01-061325	4669	0.8093	-	7471B	PPB	17 Jun 2025 11:52:46	U	SMPL
CCV54 - 1	CCV54	27631	5.0318	-	7471B	PPB	17 Jun 2025 11:55:05	U	SMPL
CCB54 - 1	CCB54	-45	-0.0576	-	7471B	PPB	17 Jun 2025 12:02:18	U	SMPL
Q2323-01 - 1	PL-01-06132025	1570	0.2394	-	7471B	PPB	17 Jun 2025 12:04:33	U	SMPL
Q2324-01 - 1	HD-01-6132025	2624	0.4332	-	7471B	PPB	17 Jun 2025 12:06:48	U	SMPL
Q2325-01 - 1	TP-8	35533	6.4850	-	7471B	PPB	17 Jun 2025 12:09:04	U	SMPL
Q2336-01 - 1	WBR-1	6337	1.1160	-	7471B	PPB	17 Jun 2025 12:11:21	U	SMPL
Q2336-03 - 1	WBR-2	1437	0.2149	-	7471B	PPB	17 Jun 2025 12:13:44	U	SMPL
Q2337-01 - 1	OR-02-061625	5728	1.0040	-	7471B	PPB	17 Jun 2025 12:16:03	U	SMPL
Q2338-01 - 1	OK-03-061625	1154	0.1629	-	7471B	PPB	17 Jun 2025 12:18:19	U	SMPL
Q2339-01 - 1	MH-C	323	0.0100	-	7471B	PPB	17 Jun 2025 12:20:38	U	SMPL
Q2340-01 - 1	TP05-MH17A-WC	4788	0.8311	-	7471B	PPB	17 Jun 2025 12:22:55	U	SMPL
Q2341-01 - 1	TP-9	91	-0.0326	-	7471B	PPB	17 Jun 2025 12:25:11	U	SMPL
CCV55 - 1	CCV55	28030	5.1052	-	7471B	PPB	17 Jun 2025 12:27:29	U	SMPL
CCB55 - 1	CCB55	-143	-0.0756	-	7471B	PPB	17 Jun 2025 12:29:44	U	SMPL
Q2341-05 - 1	EP-3	934	0.1224	-	7471B	PPB	17 Jun 2025 12:32:04	U	SMPL
Q2312-01LX5 - 1		182	-0.0159	-	7471B	PPB	17 Jun 2025 12:34:20	U	SMPL
Q2312-01A - 1		21827	3.9645	-	7471B	PPB	17 Jun 2025 12:39:13	U	SMPL
CCV56 - 1	CCV56	27494	5.0067	-	7471B	PPB	17 Jun 2025 12:41:28	U	SMPL
CCB56 - 1	CCB56	-219	-0.0896	-	7471B	PPB	17 Jun 2025 12:43:47	U	SMPL