

TOTAL SOLIDS - SM2540B

SUPERVISOR: Iwona

ANALYST: jignesh

Date: 06/13/2025

Run Number: LB136173

BalanceID: WC SC-6

OvenID: WC OVEN-1

ThermometerID: WET OVEN#1

TEMP1 IN: 104 °C 06/13/2025 11:00 **TEMP1 OUT:** 103 °C 06/13/2025 12:00
TEMP2 IN: 104 °C 06/13/2025 12:30 **TEMP2 OUT:** 103 °C 06/13/2025 13:30
TEMP3 IN: 104 °C 06/13/2025 15:30 **TEMP3 OUT:** 104 °C 06/16/2025 07:00
TEMP4 IN: 103 °C 06/16/2025 07:30 **TEMP4 OUT:** 103 °C 06/16/2025 09:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	Original weight 1st Dish+Sample weight after Drying @103-@105°C (g)	Constant weight 2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Constant weight Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB136173BL	LB136173BL	135.6871	135.6871	100	135.6871	135.6871	135.6870	0.0000	0
2	Q2301-04	WC-URBAN-FILL-C	152.1894	152.1894	100	152.1977	152.1977	152.1980	0.0083	83
3	Q2301-04DUP	WC-URBAN-FILL-CDUP	156.5840	156.5840	100	156.5920	156.5920	156.5920	0.0080	80

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

16 VB 136119 136155
VB 136173
Date: 06-13-2025 12:19:28

WorkList Name : astm ts q2301 WorkList ID : 190191 Department : Wet-Chemistry

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2301-04	WC-URBAN-FILL-C	Solid	ASTM TS	Cool 4 deg C	ENTA05	D41	06/11/2025	SM2540 B

Date/Time 06/13/25 13:00
Raw Sample Received by: SP Wey
Raw Sample Relinquished by: SP Wey

Date/Time 06/13/25 16:00
Raw Sample Received by: SP Wey
Raw Sample Relinquished by: SP Wey