

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136180

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/18/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	06/18/2025	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	06/18/2025	08:17
4	ICV	1	Water	NA	NA	20.2	7.00	06/18/2025	08:20
5	CCV1	1	Water	NA	NA	20.1	2.01	06/18/2025	08:25
6	Q2321-02	1	Water	NA	NA	23.1	7.35	06/18/2025	08:37
7	Q2321-02DUP	1	Water	NA	NA	23.2	7.36	06/18/2025	08:40
8	CCV2	1	Water	NA	NA	20.2	12.02	06/18/2025	08:44

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph q2321

WorkList ID : 190247

Department : Wet-Chemistry

Date : 06-18-2025 07:49:54

UP 16
UP 136180

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2321-02	34880	Water	pH	Cool 4 deg C	PSEG03	D51	06/13/2025	9040C

18
06/18/25 08:10

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

SA WEL
CP SM

06/18/25 10:30
CP SM
SA WEL