

Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136185

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/18/2025	10:30
2	CAL2	1	Water	NA	NA	20.2	7.00	06/18/2025	10:31
3	CAL3	1	Water	NA	NA	20.3	10.02	06/18/2025	10:33
4	ICV	1	Water	NA	NA	20.3	6.98	06/18/2025	10:37
5	CCV1	1	Water	NA	NA	20.2	2.01	06/18/2025	10:40
6	Q2344-01	1	Water	NA	NA	20.4	6.16	06/18/2025	10:47
7	Q2344-03	1	Water	NA	NA	20.3	6.86	06/18/2025	10:55
8	Q2344-05	1	Water	NA	NA	20.4	6.80	06/18/2025	11:00
9	Q2344-07	1	Water	NA	NA	20.3	6.83	06/18/2025	11:10
10	Q2344-07DUP	1	Water	NA	NA	20.4	6.84	06/18/2025	11:15
11	CCV2	1	Water	NA	NA	20.2	12.02	06/18/2025	11:20

WORKLIST(Hardcopy Internal Chain)

17196185

WorkList Name : PH Q2344 W WorkList ID : 190259 Department : Wet-Chemistry Date : 06-18-2025 10:17:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2344-01	D MW-1	Water	pH	Cool 4 deg C	LOCK01	D31	06/16/2025	SM 4500-H B
Q2344-03	D MW-3	Water	pH	Cool 4 deg C	LOCK01	D31	06/16/2025	SM 4500-H B
Q2344-05	D MW-4	Water	pH	Cool 4 deg C	LOCK01	D31	06/16/2025	SM 4500-H B
Q2344-07	P MW-2	Water	pH	Cool 4 deg C	LOCK01	D31	06/16/2025	SM 4500-H B

Date/Time 06/18/25 10:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06/18/25 13:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]