

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136192

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	83.9	06/18/2025	11:45
2	Q2307-03		1	12	100.00	>212.0	>212.0	06/18/2025	12:15
3	Q2307-03DUP		1	12	100.00	>212.0	>212.0	06/18/2025	12:45
4	Q2321-02		1	13	100.00	>212.0	>212.0	06/18/2025	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 136192

WorkList Name : FLASH POINT-061825

WorkList ID : 190258

Department : Wet-Chemistry

Date : 06-18-2025 09:45:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2307-03	LINDEN-SAA-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	D51	06/12/2025	1010B
Q2321-02	34880	Water	Flash Point	Cool 4 deg C	PSEG03	D51	06/13/2025	1010B

Date/Time 06/18/25 11:40
Raw Sample Received by: EM (WC)
Raw Sample Relinquished by: JPLAC

Date/Time 06/18/25 13:55
Raw Sample Received by: JPLAC
Raw Sample Relinquished by: EM (WC)