

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136200

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/19/2025	08:11
2	CAL2	1	Water	NA	NA	20.2	7.00	06/19/2025	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	06/19/2025	08:13
4	ICV	1	Water	NA	NA	20.2	7.00	06/19/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	06/19/2025	08:20
6	Q2351-01	1	Water	NA	NA	21.4	6.26	06/19/2025	08:30
7	Q2351-01DUP	1	Water	NA	NA	21.6	6.28	06/19/2025	08:34
8	Q2352-01	1	Water	NA	NA	22.1	6.62	06/19/2025	08:39
9	Q2354-04	1	Water	NA	NA	22.1	6.47	06/19/2025	08:45
10	Q2354-05	1	Water	NA	NA	22.3	6.85	06/19/2025	08:50
11	Q2355-05	1	Water	NA	NA	22.6	7.07	06/19/2025	09:00
12	CCV2	1	Water	NA	NA	20.2	12.02	06/19/2025	09:05

WORKLIST(Hardcopy Internal Chain)

136200

WorkList Name : ph w Q2355 WorkList ID : 190265 Department : Wet-Chemistry Date : 06-19-2025 07:40:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2351-01	FAIRLAWN-SUMP	Water	pH	Cool 4 deg C	PSEG03	D51	06/18/2025	9040C
Q2352-01	292	Water	pH	Cool 4 deg C	PSEG03	D51	06/18/2025	9040C
Q2354-04	3310	Water	pH	Cool 4 deg C	PSEG03	D51	06/18/2025	9040C
Q2354-05	3311	Water	pH	Cool 4 deg C	PSEG03	D51	06/18/2025	9040C
Q2355-05	3909	Water	pH	Cool 4 deg C	PSEG03	D51	06/18/2025	9040C

Date/Time 06-19-25 08:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06-19-25 10:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]