

LB136201


7470A

INSTRUMENT M: CV1

Linear

 μ Abs.:

40707



A= 0.0000e+000

B= 2.4425e-004

C= -1.1794e-001

Rho= 0.9995368

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/00
0.0	0.000	-0.102	-0.102	64	0.000	64	0				1
0.2	0.200	0.214	0.014	1360	0.0 %	1360					7
2.5	2.500	2.505	0.005	10737	0.0 %	10737					0
5.0	5.000	5.111	0.111	21409	0.0 %	21409					2
7.5	7.500	7.648	0.148	31795	0.0 %	31795					2
10.0	10.000	9.825	-0.175	40707	0.0 %	40707					2

LB136201 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jun 2025 09:59:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	64	-	0.0000	7470A	PPB	19 Jun 2025 10:28:09	S	Std
0.2 - 1	50.2	1360	-	0.2000	7470A	PPB	19 Jun 2025 10:30:26	S	Std
2.5 - 1	52.5	10737	-	2.5000	7470A	PPB	19 Jun 2025 10:32:43	S	Std
5.0 - 1	55	21409	-	5.0000	7470A	PPB	19 Jun 2025 10:35:00	S	Std
7.5 - 1	57.5	31795	-	7.5000	7470A	PPB	19 Jun 2025 10:37:18	S	Std
10.0 - 1	59	40707	-	10.0000	7470A	PPB	19 Jun 2025 10:50:21	S	Std
ICV45 - 1	ICV45	16587	3.9334	-	7470A	PPB	19 Jun 2025 10:53:27	U	SMPL
ICB45 - 1	ICB45	209	-0.0669	-	7470A	PPB	19 Jun 2025 10:59:49	U	SMPL
CCV63 - 1	CCV63	21229	5.0672	-	7470A	PPB	19 Jun 2025 11:02:05	U	SMPL
CCB63 - 1	CCB63	554	0.0174	-	7470A	PPB	19 Jun 2025 11:04:20	U	SMPL
CRA - 1	CRA	1271	0.1925	-	7470A	PPB	19 Jun 2025 11:06:38	U	SMPL
HighStd - 1	HighStd	41670	10.0598	-	7470A	PPB	19 Jun 2025 11:08:53	U	SMPL
ChkStd - 1	ChkStd	31942	7.6838	-	7470A	PPB	19 Jun 2025 11:14:01	U	SMPL
PB168538BL - 1	PBW	433	-0.0122	-	7470A	PPB	19 Jun 2025 11:19:03	U	SMPL
PB168538BS - 1	LCSW	14825	3.5030	-	7470A	PPB	19 Jun 2025 11:21:18	U	SMPL
Q2339-04 - 1	MH-C	487	0.0010	-	7470A	PPB	19 Jun 2025 11:27:04	U	SMPL
Q2340-04 - 1	TP05-MH17A-WC	952	0.1146	-	7470A	PPB	19 Jun 2025 11:29:20	U	SMPL
Q2341-04 - 1	TP-9	615	0.0323	-	7470A	PPB	19 Jun 2025 11:31:37	U	SMPL
Q2341-08 - 1	EP-3	610	0.0311	-	7470A	PPB	19 Jun 2025 11:33:54	U	SMPL
Q2347-04 - 1	TP-10-9	340	-0.0349	-	7470A	PPB	19 Jun 2025 11:36:10	U	SMPL
CCV64 - 1	CCV64	21379	5.1038	-	7470A	PPB	19 Jun 2025 11:38:25	U	SMPL
CCB64 - 1	CCB64	94	-0.0950	-	7470A	PPB	19 Jun 2025 11:44:11	U	SMPL
Q2347-04DUP - 1	TP-10-9DUP	540	0.0140	-	7470A	PPB	19 Jun 2025 11:46:26	U	SMPL
Q2347-04MS - 1	TP-10-9MS	15211	3.5973	-	7470A	PPB	19 Jun 2025 11:48:41	U	SMPL
Q2347-04MSD - 1	TP-10-9MSD	15179	3.5895	-	7470A	PPB	19 Jun 2025 11:50:57	U	SMPL
PB168514TB - 1	PB168514TB	293	-0.0464	-	7470A	PPB	19 Jun 2025 11:55:20	U	SMPL
Q2347-04LX5 - 1		80	-0.0984	-	7470A	PPB	19 Jun 2025 11:57:36	U	SMPL
Q2347-04A - 1		16204	3.8398	-	7470A	PPB	19 Jun 2025 11:59:52	U	SMPL
CCV65 - 1	CCV65	21364	5.1001	-	7470A	PPB	19 Jun 2025 12:02:08	U	SMPL
CCB65 - 1	CCB65	158	-0.0793	-	7470A	PPB	19 Jun 2025 12:04:27	U	SMPL