



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/20/2025

OVENTEMP IN Celsius(°C): 108
Time IN: 16:50
In Date: 06/19/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 06/20/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136203

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2362-01	TP-11	1	1.15	10.82	11.97	10.9	90.1	
Q2362-02	TP-11-EPH	2	1.18	9.88	11.06	9.94	88.7	
Q2362-03	TP-11-VOC	3	1.19	10.70	11.89	10.74	89.3	
Q2362-04	TP-11	4	1.15	10.82	11.97	10.9	90.1	
Q2362-05	EP-6	5	1.16	10.57	11.73	10.57	89.0	
Q2362-06	EP-6-EPH	6	1.16	10.34	11.5	10.28	88.2	
Q2362-07	EP-6-VOC	7	1.18	10.32	11.5	10.38	89.1	
Q2363-01	GCAP1	8	1.19	9.95	11.14	9.7	85.5	
Q2364-01	GAV1A	9	1.15	10.10	11.25	9.21	79.8	
Q2364-02	GAV1B	10	1.18	10.78	11.96	10.4	85.5	
Q2364-03	GAV2A	11	1.19	10.06	11.25	9.47	82.3	
Q2364-04	GAV2B	12	1.19	10.53	11.72	9.26	76.6	
Q2365-01	BU-703-COMP-01	13	1.15	10.74	11.89	10.68	88.7	
Q2365-02	BU-703-COMP-02	14	1.14	10.60	11.74	9.59	79.7	
Q2365-03	BU-703-COMP-03	15	1.17	10.22	11.39	9.59	82.4	
Q2365-04	BU-703-COMP-04	16	1.15	10.58	11.73	10.55	88.8	
Q2365-05	BU-703-COMP-05	17	1.13	10.19	11.32	8.67	74.0	
Q2365-06	BU-703-COMP-06	18	1.15	10.59	11.74	10.11	84.6	
Q2365-07	BU-703-COMP-07	19	1.13	10.28	11.41	10.02	86.5	
Q2365-08	BU-703-COMP-08	20	1.13	10.20	11.33	10.28	89.7	
Q2365-09	BU-703-COMP-09	21	1.14	10.54	11.68	9.73	81.5	
Q2367-01	EP-4	22	1.18	10.29	11.47	10.58	91.4	
Q2367-02	EP-4-EPH	23	1.19	10.01	11.2	10.24	90.4	
Q2367-03	EP-4-VOC	24	1.13	9.84	10.97	10.03	90.4	
Q2367-05	EP-5	25	1.17	10.39	11.56	10.44	89.2	
Q2367-06	EP-5-EPH	26	1.18	10.54	11.72	10.65	89.8	
Q2367-07	EP-5-VOV	27	1.15	10.45	11.6	10.56	90.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$