

LB136204


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

62635



A= 0.0000e+000

B= 1.6042e-004

C= 4.4623e-002

Rho= 0.9997504

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.071	0.071	164	0.000	164					4.20
0.2	0.200	0.225	0.025	1124	0.0 %	1124					13
2.5	2.500	2.492	-0.008	15259	0.0 %	15259					0
5.0	5.000	4.838	-0.162	29882	0.0 %	29882					~3
7.5	7.500	7.481	-0.019	46356	0.0 %	46356					0
10.0	10.000	10.092	0.092	62635	0.0 %	62635					5

LB136204 INSTRUMENT ID :CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jun 2025 13:48:22

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	164	-	0.0000	7471B	PPB	19 Jun 2025 13:55:04	S	Std
0.2 - 1	50.2	1124	-	0.2000	7471B	PPB	19 Jun 2025 13:57:22	S	Std
2.5 - 1	52.5	15259	-	2.5000	7471B	PPB	19 Jun 2025 13:59:38	S	Std
5.0 - 1	55	29882	-	5.0000	7471B	PPB	19 Jun 2025 14:01:56	S	Std
7.5 - 1	57.5	46356	-	7.5000	7471B	PPB	19 Jun 2025 14:04:14	S	Std
10.0 - 1	59.5	62635	-	10.0000	7471B	PPB	19 Jun 2025 14:06:33	S	Std
ICV46 - 1	ICV46	23925	3.8826	-	7471B	PPB	19 Jun 2025 14:12:11	U	SMPL
ICB46 - 1	ICB46	122	0.0642	-	7471B	PPB	19 Jun 2025 14:17:15	U	SMPL
CCV66 - 1	CCV66	29868	4.8360	-	7471B	PPB	19 Jun 2025 14:19:30	U	SMPL
CCB66 - 1	CCB66	-149	0.0207	-	7471B	PPB	19 Jun 2025 14:21:45	U	SMPL
CRA - 1	CRA	1105	0.2219	-	7471B	PPB	19 Jun 2025 14:24:04	U	SMPL
HighStd - 1	HighStd	62307	10.0398	-	7471B	PPB	19 Jun 2025 14:26:19	U	SMPL
ChkStd - 1	ChkStd	40702	6.5740	-	7471B	PPB	19 Jun 2025 14:28:34	U	SMPL
PB168542BL - 1	PBS	127	0.0650	-	7471B	PPB	19 Jun 2025 14:33:59	U	SMPL
PB168542BS - 1	LCSS	23448	3.8061	-	7471B	PPB	19 Jun 2025 14:38:24	U	SMPL
Q2347-01 - 1	TP-10	1854	0.3420	-	7471B	PPB	19 Jun 2025 14:50:37	U	SMPL
Q2347-01DUP - 1	TP-10DUP	1554	0.2939	-	7471B	PPB	19 Jun 2025 14:52:54	U	SMPL
Q2347-01MS - 1	TP-10MS	23385	3.7960	-	7471B	PPB	19 Jun 2025 14:57:51	U	SMPL
Q2347-01MSD - 1	TP-10MSD	22675	3.6821	-	7471B	PPB	19 Jun 2025 15:00:07	U	SMPL
CCV67 - 1	CCV67	29544	4.7840	-	7471B	PPB	19 Jun 2025 15:02:25	U	SMPL
CCB67 - 1	CCB67	-254	0.0039	-	7471B	PPB	19 Jun 2025 15:04:43	U	SMPL
Q2354-01 - 1	3321	1549	0.2931	-	7471B	PPB	19 Jun 2025 15:07:02	U	SMPL
Q2355-01 - 1	3897	933	0.1943	-	7471B	PPB	19 Jun 2025 15:09:17	U	SMPL
Q2355-03 - 1	20071	1093	0.2200	-	7471B	PPB	19 Jun 2025 15:11:32	U	SMPL
Q2357-01 - 1	TR-06-06182025	2170	0.3927	-	7471B	PPB	19 Jun 2025 15:13:48	U	SMPL
Q2358-01 - 1	NB-07-06182025	1188	0.2352	-	7471B	PPB	19 Jun 2025 15:16:04	U	SMPL
Q2347-01LX5 - 1		228	0.0812	-	7471B	PPB	19 Jun 2025 15:18:21	U	SMPL
Q2347-01A - 1		21863	3.5519	-	7471B	PPB	19 Jun 2025 15:20:37	U	SMPL
Q2126-01 - 1	MDL-SOIL	93	0.0595	-	7471B	PPB	19 Jun 2025 15:22:54	U	SMPL
Q2126-01RE - 1	MDL-SOILRE	649	0.1487	-	7471B	PPB	19 Jun 2025 15:25:15	U	SMPL
Q2126-03 - 1	MDL-SOIL-03-QT2-2025	77	0.0570	-	7471B	PPB	19 Jun 2025 15:32:34	U	SMPL
CCV68 - 1	CCV68	27887	4.5182	-	7471B	PPB	19 Jun 2025 15:36:32	U	SMPL
CCB68 - 1	CCB68	-220	0.0093	-	7471B	PPB	19 Jun 2025 15:38:47	U	SMPL