

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB136206

Reagent/Standard	Lot/Log #
COD calibration std. 150 ppm	WP113238
COD calibration std. 100 ppm	WP113237
COD calibration std. 50 ppm	WP113235
COD calibration std. 10 ppm	WP113234
COD calibration std. 0 ppm	WP113233
COD ICV-LCS std, 50ppm	WP113240
COD calibration std. 75 ppm	WP113236
COD Digestion Vials Low Level 0-150Mg/L	W3129
COD CCV std, 50ppm	WP113586
COD ICV-LCS std, 50ppm	WP113587
RL CHECK	WP113588

Temp In(C): 148	Date In: 06/19/2025	Time In: 09:40
Temp Out(C): 151	Date Out: 06/19/2025	Time Out: 11:40

Intercept: 0.8179

Slope: 0.9847

Regression: 0.9995

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.831		05/28/2025	13:10
2	CAL2	10	1	Water	9.000	8.309	-16.9	05/28/2025	13:10
3	CAL3	50	1	Water	52.000	51.977	4	05/28/2025	13:11
4	CAL4	75	1	Water	77.000	77.366	3.2	05/28/2025	13:11
5	CAL5	100	1	Water	99.000	99.708	-0.3	05/28/2025	13:12
6	CAL6	150	1	Water	147.000	148.453	-1	05/28/2025	13:12

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB136206

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	50.962	05/28/2025	13:13
2	ICB		NA	NA	1	Water	0.000	-0.831	05/28/2025	13:13
3	CCV1	50	NA	NA	1	Water	48.000	47.915	06/19/2025	13:30
4	CCB1		NA	NA	1	Water	1.000	0.185	06/19/2025	13:30
5	RL Check	10	NA	NA	1	Water	8.000	7.294	06/19/2025	13:31
6	LB136206BL		NA	NA	1	Water	1.000	0.185	06/19/2025	13:31
7	LB136206BS	50	NA	NA	1	Water	51.000	50.962	06/19/2025	13:32
8	Q2334-01		NA	NA	1	Water	15.000	14.402	06/19/2025	13:32
9	Q2334-01DUP		NA	NA	1	Water	16.000	15.418	06/19/2025	13:33
10	Q2334-01MS	50	NA	NA	1	Water	61.000	61.117	06/19/2025	13:33
11	Q2334-01MSD	50	NA	NA	1	Water	60.000	60.102	06/19/2025	13:34
12	Q2334-02		NA	NA	1	Water	16.000	15.418	06/19/2025	13:34
13	Q2334-03		NA	NA	1	Water	53.000	52.993	06/19/2025	13:35
14	CCV2	50	NA	NA	1	Water	49.000	48.931	06/19/2025	13:35
15	CCB2		NA	NA	1	Water	1.000	0.185	06/19/2025	13:36

WORKLIST(Hardcopy Internal Chain)

LB136206

WorkList Name : COD-061625

WorkList ID : 190215

Department : Wet-Chemistry

Date : 06-16-2025 09:46:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2334-01	MW3	Water	COD	Conc H2SO4 to pH < 2	GENV01	D52	06/12/2025	SM5220 D
Q2334-02	MW4	Water	COD	Conc H2SO4 to pH < 2	GENV01	D52	06/12/2025	SM5220 D
Q2334-03	GBTW1	Water	COD	Conc H2SO4 to pH < 2	GENV01	D52	06/12/2025	SM5220 D

Date/Time 06/19/25 09:15
 Raw Sample Received by: 12(50)
 Raw Sample Relinquished by: 20(00)

Date/Time 06/19/25 10:05
 Raw Sample Received by: 20(00)
 Raw Sample Relinquished by: 12(50)