

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136224

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/23/2025	08:17
2	CAL2	1	Water	NA	NA	20.2	7.00	06/23/2025	08:18
3	CAL3	1	Water	NA	NA	20.3	10.02	06/23/2025	08:20
4	ICV	1	Water	NA	NA	20.3	7.02	06/23/2025	08:22
5	CCV1	1	Water	NA	NA	20.3	2.01	06/23/2025	08:25
6	Q2382-01	1	Water	NA	NA	24.1	7.45	06/23/2025	08:37
7	Q2382-01DUP	1	Water	NA	NA	24.2	7.46	06/23/2025	08:39
8	Q2385-01	1	Water	NA	NA	22.6	7.27	06/23/2025	08:45
9	Q2391-02	1	Water	NA	NA	24.4	7.60	06/23/2025	08:50
10	CCV2	1	Water	NA	NA	20.2	12.02	06/23/2025	09:00

WORKLIST(Hardcopy Internal Chain)

✓ 136224

WorkList Name : ph q2385 w WorkList ID : 190307 Department : Wet-Chemistry Date : 06-23-2025 07:58:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2382-01	B FMI109	Water	pH	Cool 4 deg C	PSEG03	D51	06/20/2025	9040C
Q2385-01	F A5311	Water	pH	Cool 4 deg C	PSEG03	D51	06/20/2025	9040C
Q2391-02	A AUD-1600	Water	pH	Cool 4 deg C	PSEG03	D41	06/20/2025	9040C

Date/Time 06/23/25 08:10
Raw Sample Received by: TB (WLC)
Raw Sample Relinquished by: CP SM

Date/Time 06/23/25 11:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: SP (WLC)