

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB136231

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	84.2	06/23/2025	12:30
2	Q2354-03		1	4	7.00	44.6	44.6	06/23/2025	13:00
3	Q2354-03DUP		1	4	7.00	44.6	44.6	06/23/2025	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136231

WorkList Name : FP-062025-02

WorkList ID : 190304

Department : Wet-Chemistry

Date : 06-20-2025 15:02:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2354-03	3323	Solid	Flash Point	Cool 4 deg C	PSEG03	D51	06/18/2025	1010B

Date/Time 06/23/25 12:10
 Raw Sample Received by: 12(-C)
 Raw Sample Relinquished by: jeff

Date/Time 06/23/25 14:15
 Raw Sample Received by: jeff
 Raw Sample Relinquished by: 12(-C)