

LB136241


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

55133



A= 0.0000e+000

B= 1.8213e-004

C= -1.7318e-002

Rho= 0.9998973

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.022	-0.022	-27	0.000	-27					13
0.2	0.200	0.175	-0.025	1058	0.0 %	1058					2
2.5	2.500	2.555	0.055	14122	0.0 %	14122					1
5.0	5.000	5.060	0.060	27877	0.0 %	27877					1
7.5	7.500	7.408	-0.092	40768	0.0 %	40768					0
10.0	10.000	10.024	0.024	55133	0.0 %	55133					

LB136241 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jun 2025 11:07:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-27	-	0.0000	7471B	PPB	24 Jun 2025 11:16:02	S	Std
0.2 - 1	50.3	1058	-	0.2000	7471B	PPB	24 Jun 2025 11:18:19	S	Std
2.5 - 1	52.5	14122	-	2.5000	7471B	PPB	24 Jun 2025 11:20:36	S	Std
5.0 - 1	55	27877	-	5.0000	7471B	PPB	24 Jun 2025 11:25:39	S	Std
7.5 - 1	57.5	40768	-	7.5000	7471B	PPB	24 Jun 2025 11:27:56	S	Std
10.0 - 1	59	55133	-	10.0000	7471B	PPB	24 Jun 2025 11:30:15	S	Std
ICV02 - 1	ICV02	21399	3.8801	-	7471B	PPB	24 Jun 2025 11:33:23	U	SMPL
ICB02 - 1	ICB02	-196	-0.0530	-	7471B	PPB	24 Jun 2025 11:35:38	U	SMPL
CCV05 - 1	CCV05	26461	4.8021	-	7471B	PPB	24 Jun 2025 11:37:56	U	SMPL
CCB05 - 1	CCB05	-261	-0.0649	-	7471B	PPB	24 Jun 2025 11:40:11	U	SMPL
CRA - 1	CRA	1259	0.2120	-	7471B	PPB	24 Jun 2025 11:46:33	U	SMPL
HighStd - 1	HighStd	65439	11.9013	-	7471B	PPB	24 Jun 2025 11:48:48	U	SMPL
ChkStd - 1	ChkStd	40925	7.4365	-	7471B	PPB	24 Jun 2025 11:51:03	U	SMPL
PB168589BL - 1	PBS	-239	-0.0608	-	7471B	PPB	24 Jun 2025 11:53:29	U	SMPL
PB168589BS - 1	LCSS	23247	4.2167	-	7471B	PPB	24 Jun 2025 11:55:52	U	SMPL
Q2380-01 - 1	72-11948	2900	0.5109	-	7471B	PPB	24 Jun 2025 11:58:08	U	SMPL
Q2380-03 - 1	VNJ-211	145564	26.4947	-	7471B	PPB	24 Jun 2025 12:00:28	U	SMPL
Q2380-05 - 1	RBR-200059	1829	0.3158	-	7471B	PPB	24 Jun 2025 12:02:47	U	SMPL
Q2381-01 - 1	291431	107598	19.5798	-	7471B	PPB	24 Jun 2025 12:05:28	U	SMPL
Q2381-03 - 1	VNJ-207	13919	2.5178	-	7471B	PPB	24 Jun 2025 12:07:44	U	SMPL
CCV06 - 1	CCV06	28275	5.1325	-	7471B	PPB	24 Jun 2025 12:13:06	U	SMPL
CCB06 - 1	CCB06	-264	-0.0654	-	7471B	PPB	24 Jun 2025 12:15:22	U	SMPL
Q2381-05 - 1	72-11929	5136	0.9181	-	7471B	PPB	24 Jun 2025 12:17:40	U	SMPL
Q2383-01 - 1	RT-5376	119	0.0044	-	7471B	PPB	24 Jun 2025 12:19:55	U	SMPL
Q2384-01 - 1	OK-01-6202025	1333	0.2255	-	7471B	PPB	24 Jun 2025 12:22:12	U	SMPL
Q2386-01 - 1	SU-1-062025	247	0.0277	-	7471B	PPB	24 Jun 2025 12:24:29	U	SMPL
Q2388-01 - 1	TP-12	139	0.0080	-	7471B	PPB	24 Jun 2025 12:26:46	U	SMPL
Q2389-01 - 1	MH-J-I	374	0.0508	-	7471B	PPB	24 Jun 2025 12:44:30	U	SMPL
Q2389-01DUP - 1	MH-J-IDUP	358	0.0479	-	7471B	PPB	24 Jun 2025 12:46:46	U	SMPL
Q2389-01MS - 1	MH-J-IMS	23908	4.3371	-	7471B	PPB	24 Jun 2025 12:49:03	U	SMPL
Q2389-01MSD - 1	MH-J-IMSD	23933	4.3417	-	7471B	PPB	24 Jun 2025 12:51:20	U	SMPL
CCV07 - 1	CCV07	27714	5.0303	-	7471B	PPB	24 Jun 2025 12:56:27	U	SMPL
CCB07 - 1	CCB07	-254	-0.0636	-	7471B	PPB	24 Jun 2025 12:58:42	U	SMPL
Q2389-01LX5 - 1		114	0.0034	-	7471B	PPB	24 Jun 2025 13:01:01	U	SMPL
Q2389-01A - 1		22804	4.1360	-	7471B	PPB	24 Jun 2025 13:03:17	U	SMPL
Q2380-03DLX5 - 1		28481	5.1700	-	7471B	PPB	24 Jun 2025 13:10:54	U	SMPL
Q2381-01DLX5 - 1		21914	3.9739	-	7471B	PPB	24 Jun 2025 13:15:49	U	SMPL
CCV08 - 1	CCV08	25034	4.5422	-	7471B	PPB	24 Jun 2025 13:18:06	U	SMPL
CCB08 - 1	CCB08	-277	-0.0678	-	7471B	PPB	24 Jun 2025 13:20:25	U	SMPL