

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136250

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/24/2025	16:00
2	CAL2	1	Water	NA	NA	20.2	4.01	06/24/2025	16:03
3	CAL3	1	Water	NA	NA	20.2	10.02	06/24/2025	16:05
4	ICV	1	Water	NA	NA	20.2	7.00	06/24/2025	16:10
5	CCV1	1	Water	NA	NA	20.3	2.01	06/24/2025	16:11
6	Q2403-01	1	Water	NA	NA	23.2	7.44	06/24/2025	16:15
7	Q2403-01DUP	1	Water	NA	NA	23.4	7.45	06/24/2025	16:20
8	CCV2	1	Water	NA	NA	20.2	12.02	06/24/2025	16:25

WORKLIST(Hardcopy Internal Chain)

NP136250

WorkList Name : PH W Q2403

WorkList ID : 190362

Department : Wet-Chemistry

Date : 06-24-2025 15:27:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2403-01	B LAW-25-0092	Water	pH	Cool 4 deg C	PSEG03	A32	06/23/2025	9040C

Date/Time 06/24/25 15:35

Raw Sample Received by: SA WOC

Raw Sample Relinquished by: SA WOC

Date/Time 06/24/25

Raw Sample Received by: SA WOC

Raw Sample Relinquished by: SA WOC

18:00

SA WOC

SA WOC