

Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136252

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP113655
Turbidity Calibration std, 0NTU	WP113654
Turbidity Calibration std, 20NTU	WP113649
Turbidity Calibration std, 40NTU	WP113648
Turbidity Calibration std, 1NTU	WP113652
Turbidity Calibration std, 5NTU	WP113651
Turbidity Calibration - CCV std, 10 NTU	WP113650
10 NTU Standard 500 ml	W3116

Intercept: -0.1019

Slope: 1.0023

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.102	0.20		06/24/2025	13:20
2	CAL2	1	1	0.962	1.06	6.1	06/24/2025	13:23
3	CAL3	5	1	4.960	5.05	1	06/24/2025	13:26
4	CAL4	10	1	9.740	9.82	-1.8	06/24/2025	13:30
5	CAL5	20	1	19.600	19.66	-1.7	06/24/2025	13:33
6	CAL6	40	1	40.200	40.21	0.5	06/24/2025	13:36



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136252

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.650	9.730	06/24/2025	13:40
2	ICB		1	0.112	0.213	06/24/2025	13:43
3	CCV1	10	1	10.400	10.478	06/24/2025	13:46
4	CCB1		1	0.105	0.206	06/24/2025	13:50
5	LB136252BL		1	0.098	0.199	06/24/2025	13:53
6	Q2126-12		1	0.560	0.660	06/24/2025	13:56
7	CCV2	10	1	9.660	9.739	06/24/2025	14:00
8	CCB2		1	0.085	0.186	06/24/2025	14:03