

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136256

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	06/25/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	06/25/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	06/25/2025	08:12
4	ICV	1	Water	NA	NA	20.2	7.01	06/25/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	06/25/2025	08:20
6	Q2393-01	1	Solid	20.02	20	24.4	6.08	06/25/2025	08:30
7	Q2393-01DUP	1	Solid	20.03	20	24.6	6.10	06/25/2025	08:34
8	Q2393-03	1	Solid	20.02	20	24.7	5.60	06/25/2025	08:40
9	Q2393-05	1	Solid	20.03	20	25.3	9.75	06/25/2025	08:50
10	Q2393-07	1	Solid	20.04	20	24.9	10.06	06/25/2025	09:10
11	Q2393-09	1	Solid	20.02	20	25.0	9.46	06/25/2025	09:15
12	Q2393-11	1	Solid	20.03	20	24.6	9.77	06/25/2025	09:25
13	Q2410-01	1	Solid	20.02	20	24.6	7.34	06/25/2025	09:35
14	CCV2	1	Water	NA	NA	20.3	12.02	06/25/2025	09:40

WORKLIST(Hardcopy Internal Chain)

136256

WorkList Name : ph s q2393

WorkList ID : 190363

Department : Wet-Chemistry

Date : 06-25-2025 07:49:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2393-01	PARK AVE -1	Solid	pH	Cool 4 deg C	UNIO02	A22	06/23/2025	9045D
Q2393-03	PARK AVE -2	Solid	pH	Cool 4 deg C	UNIO02	A22	06/23/2025	9045D
Q2393-05	PARK AVE -3	Solid	pH	Cool 4 deg C	UNIO02	A22	06/23/2025	9045D
Q2393-07	PARK AVE -4	Solid	pH	Cool 4 deg C	UNIO02	A22	06/23/2025	9045D
Q2393-09	PARK AVE -5	Solid	pH	Cool 4 deg C	UNIO02	A22	06/23/2025	9045D
Q2393-11	PARK AVE -6	Solid	pH	Cool 4 deg C	UNIO02	A22	06/23/2025	9045D
Q2410-01	TRE-25-0021	Solid	pH	Cool 4 deg C	PSEG03	D41	06/24/2025	9045D

Date/Time 06/25/25 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/25/25 11:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]