



Analysis Method: 9034 ANALYST: rubina

Parameter: Sulfide SUPERVISOR REVIEW BY: Iwona

 Run Number:
 LB136272
 Constant:
 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #			
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105			
IODINE SOLUTION .025N 1L	W3213			
Starch Solution, 4L	W3149			

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB168395BL		1	50	50	5.00	0.00	4.92	4.92	0.08	0.00	0.00	06/25/2025	17:00
2	PB168395BS	25	1	50	50	5.00	0.00	1.96	1.96	3.04	2.96	23.68	06/25/2025	17:03
3	Q2126-09		1	50	50	5.00	0.00	4.86	4.86	0.14	0.06	0.48	06/25/2025	17:06

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 \* Normality1) - ((T1 - Value Corrected With Blank) \* Normality2)) \* Constant / Initial Volume