

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136280

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/26/2025	08:22
2	CAL2	1	Water	NA	NA	20.2	7.00	06/26/2025	08:23
3	CAL3	1	Water	NA	NA	20.2	10.02	06/26/2025	08:25
4	ICV	1	Water	NA	NA	20.5	7.03	06/26/2025	08:33
5	CCV1	1	Water	NA	NA	20.2	2.01	06/26/2025	08:37
6	Q2418-01	1	Water	NA	NA	21.1	7.50	06/26/2025	08:45
7	Q2418-01DUP	1	Water	NA	NA	21.2	7.51	06/26/2025	08:48
8	CCV2	1	Water	NA	NA	20.2	12.02	06/26/2025	08:50

WORKLIST(Hardcopy Internal Chain)

VB 136280

WorkList Name : ph q2418

WorkList ID : 190389

Department : Wet-Chemistry

Date : 06-26-2025 07:47:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2418-01	P VAC-TRUCK-4074	Water	pH	Cool 4 deg C	PSEG03	D41	06/25/2025	9040C

Date/Time 06/26/25 08:00
Raw Sample Received by: SA WCI
Raw Sample Relinquished by: CP SM

Date/Time 06/26/25 11:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: SA WCI