

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136289

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/27/2025	08:30
2	Q2422-01		1	11	100.00	>212.0	>212.0	06/27/2025	09:00
3	Q2422-01DUP		1	11	100.00	>212.0	>212.0	06/27/2025	09:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

1B136289

WorkList Name : FP SOIL-6-25 WorkList ID : 190376 Department : Wet-Chemistry Date : 06-25-2025 13:39:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2422-01	60263	Solid	Flash Point	Cool 4 deg C	PSEG03	D51	06/25/2025	1010B

Date/Time 06/27/25 08:20
Raw Sample Received by: EM(wc)
Raw Sample Relinquished by: JN(wc)

Date/Time 06/27/25 10:10
Raw Sample Received by: JN(wc)
Raw Sample Relinquished by: EM(wc)