

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136290

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	83.9	06/26/2025	13:10
2	Q2403-01		1	12	100.00	>212.0	>212.0	06/26/2025	13:40
3	Q2403-01DUP		1	12	100.00	>212.0	>212.0	06/26/2025	14:10
4	Q2418-01		1	11	100.00	>212.0	>212.0	06/26/2025	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

13136290

WorkList Name : FP 6-25

WorkList ID : 190384

Department : Wet-Chemistry

Date : 06-25-2025 17:21:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2403-01	LAW-25-0092	Water	Flash Point	Cool 4 deg C	PSEG03	A32	06/23/2025	1010B
Q2418-01	VAC-TRUCK-4074	Water	Flash Point	Cool 4 deg C	PSEG03	D41	06/25/2025	1010B

Date/Time 06/26/25 13:00
 Raw Sample Received by: EM(WC)
 Raw Sample Relinquished by: JS(OO)

Date/Time 06/26/25 15:20
 Raw Sample Received by: JS(OO)
 Raw Sample Relinquished by: EM(WC)