

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB136299

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP113735
Residual chlorine ICV-LCS, 0.4PPM	WP113734
Chlorine Calibration std, 0.1ppm	WP113729
Chlorine Calibration std, 0.2ppm	WP113730
Chlorine Calibration std, 0.8ppm	WP113731
Chlorine Calibration std, 0.0ppm	WP113728
Chlorine Calibration std, 1.6ppm	WP113732
Residual Chlorine Calibration and CCV std, 0	WP113733
Total Chlorine Powder Pillows	W3147

Intercept: -0.0083

Slope: 0.9580

Regression: 0.999637

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.01		06/26/2025	14:00
2	CAL2	0.1	1	0.000	0.090	0.090	0.10	3	06/26/2025	14:03
3	CAL3	0.2	1	0.000	0.180	0.180	0.20	-1.5	06/26/2025	14:06
4	CAL4	0.4	1	0.000	0.350	0.350	0.37	-6.5	06/26/2025	14:09
5	CAL5	0.8	1	0.000	0.780	0.780	0.82	2.9	06/26/2025	14:12
6	CAL6	1.6	1	0.000	1.520	1.520	1.60	-0.3	06/26/2025	14:15

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB136299

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3400	0.3400	0.3640	06/26/2025	14:18
2	ICB				1	0.0000	0.0000	0.0000	0.0090	06/26/2025	14:21
3	CCV1			0.4	1	0.0000	0.3500	0.3500	0.3740	06/26/2025	14:24
4	CCB1				1	0.0000	0.0000	0.0000	0.0090	06/26/2025	14:27
5	LB136299BL	50	50		1	0.0000	0.0000	0.0000	0.0090	06/26/2025	14:30
6	LB136299BS	50	50	0.4	1	0.0000	0.3600	0.3600	0.3840	06/26/2025	14:33
7	Q2126-09	50	50		1	0.0000	0.0500	0.0500	0.0610	06/26/2025	14:36
8	CCV2			0.4	1	0.0000	0.3600	0.3600	0.3840	06/26/2025	14:39
9	CCB2				1	0.0000	0.0000	0.0000	0.0090	06/26/2025	14:42