

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136303

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/26/2025	16:10
2	CAL2	1	Water	NA	NA	20.3	7.01	06/26/2025	16:11
3	CAL3	1	Water	NA	NA	20.3	10.02	06/26/2025	16:15
4	ICV	1	Water	NA	NA	20.2	7.00	06/26/2025	16:20
5	CCV1	1	Water	NA	NA	20.3	2.01	06/26/2025	16:27
6	Q2428-01	1	Water	NA	NA	24.7	7.07	06/26/2025	16:40
7	Q2428-01DUP	1	Water	NA	NA	24.9	7.09	06/26/2025	16:44
8	CCV2	1	Water	NA	NA	20.2	12.02	06/26/2025	16:50