

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136308

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	29.00	84.2	83.5	06/27/2025	12:30
2	Q2428-01		1	15	100.00	>212.0	>212.0	06/27/2025	13:00
3	Q2428-01DUP		1	15	100.00	>212.0	>212.0	06/27/2025	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

1B136308

WorkList Name : FP-062725

WorkList ID : 190442

Department : Wet-Chemistry

Date : 06-27-2025 10:01:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2428-01	A4212	Water	Flash Point	Cool 4 deg C	PSEG03	D41	06/26/2025	1010B

Date/Time 6/27/25 12:20
 Raw Sample Received by: EM (WSC)
 Raw Sample Relinquished by: [Signature]

Date/Time 6/27/25 14:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: EM (WSC)