

L3136313


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

36979



A= 0.0000e+000

B= 2.7097e-004

C= 1.8680e-002

Rho= 0.9999310

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%d
0.0	0.000	-0.007	-0.007	-96	0.000	-96					
0.05	0.050	0.060	0.010	153	0.0 %	153	0				20
0.2	0.200	0.207	0.007	695	0.0 %	695					3
2.5	2.500	2.478	-0.022	9075	0.0 %	9075					-1
5.0	5.000	5.060	0.060	18605	0.0 %	18605					1
7.5	7.500	7.413	-0.087	27290	0.0 %	27290					-1
10.0	10.000	10.039	0.039	36979	0.0 %	36979					0

LB136313 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Jun 2025 12:11:10

Type	Sample ID	Extended ID	Conc	μ Abs	Units	Date/Method	Type	Std Conc
S	0.0 - 1	SC	-	-96	PPB	27 Jun 2025 12:23:21245.1	Std	0.0000
S	0.05 - 1	SC	-	153	PPB	27 Jun 2025 12:25:38245.1	Std	0.0500
S	0.2 - 1	SC	-	695	PPB	27 Jun 2025 12:27:55245.1	Std	0.2000
S	2.5 - 1	SC	-	9075	PPB	27 Jun 2025 12:30:12245.1	Std	2.5000
S	5.0 - 1	SC	-	18605	PPB	27 Jun 2025 12:32:28245.1	Std	5.0000
S	7.5 - 1	SC	-	27290	PPB	27 Jun 2025 12:34:45245.1	Std	7.5000
S	10.0 - 1	SC	-	36979	PPB	27 Jun 2025 12:48:32245.1	Std	10.0000
U	ICV09 - 1	ICV09	4.1857	15378	PPB	27 Jun 2025 12:54:54245.1	SMPL	-
U	ICB09 - 1	ICB09	-0.0423	-225	PPB	27 Jun 2025 12:57:09245.1	SMPL	-
U	CCV27 - 1	CCV27	5.1503	18938	PPB	27 Jun 2025 12:59:26245.1	SMPL	-
U	CCB27 - 1	CCB27	-0.0217	-149	PPB	27 Jun 2025 13:01:41245.1	SMPL	-
U	CRA - 1	CRA	0.0466	103	PPB	27 Jun 2025 13:06:58245.1	SMPL	-
U	HighStd - 1	HighStd	9.9403	36615	PPB	27 Jun 2025 13:09:12245.1	SMPL	-
U	ChkStd - 1	ChkStd	7.3604	27094	PPB	27 Jun 2025 13:11:27245.1	SMPL	-
U	PB168650BL - 1	PBW	0.0006	-67	PPB	27 Jun 2025 13:16:44245.1	SMPL	-
U	PB168650BS - 1	LCSW	3.8927	14297	PPB	27 Jun 2025 13:22:19245.1	SMPL	-
U	Q2126-10 - 1	LOD-MDL-WATER-04-QT2-2025	0.0428	89	PPB	27 Jun 2025 13:34:04245.1	SMPL	-
U	Q2126-11 - 1	LOQ-WATER-05-QT2-2025	0.2103	707	PPB	27 Jun 2025 13:39:15245.1	SMPL	-
U	Q2376-02 - 1	EFFLUENT-METALS	0.1647	539	PPB	27 Jun 2025 13:43:07245.1	SMPL	-
U	Q2376-02DUP - 1	EFFLUENT-METALSDUP	0.1046	317	PPB	27 Jun 2025 13:45:26245.1	SMPL	-
U	Q2376-02MS - 1	EFFLUENT-METALSMS	3.8896	14286	PPB	27 Jun 2025 13:47:42245.1	SMPL	-
U	CCV28 - 1	CCV28	4.9682	18266	PPB	27 Jun 2025 13:49:57245.1	SMPL	-
U	CCB28 - 1	CCB28	-0.0553	-273	PPB	27 Jun 2025 13:52:15245.1	SMPL	-
U	Q2376-02MSD - 1	EFFLUENT-METALSMSD	3.7684	13838	PPB	27 Jun 2025 13:54:32245.1	SMPL	-
U	Q2376-02LX5 - 1		-0.0133	-118	PPB	27 Jun 2025 13:56:47245.1	SMPL	-
U	Q2376-02A - 1		4.7043	17292	PPB	27 Jun 2025 13:59:05245.1	SMPL	-
U	CCV29 - 1	CCV29	4.8840	17955	PPB	27 Jun 2025 14:01:21245.1	SMPL	-
U	CCB29 - 1	CCB29	-0.0564	-277	PPB	27 Jun 2025 14:03:39245.1	SMPL	-