

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136328

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	06/30/2025	12:10
2	Q2446-01		1	14	100.00	>212.0	>212.0	06/30/2025	12:40
3	Q2446-01DUP		1	14	100.00	>212.0	>212.0	06/30/2025	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6136328

WorkList Name : fp-6-30

WorkList ID : 190453

Department : Wet-Chemistry

Date : 06-30-2025 08:12:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2446-01	MR-BUR-LNG-19	Solid	Flash Point	Cool 4 deg C	PSEG03	D41	06/27/2025	1010B

Date/Time 06/30/2025 12:00
Raw Sample Received by: RM (wc)
Raw Sample Relinquished by: RM (wc)

Date/Time 06/30/2025 13:50
Raw Sample Received by: RM (wc)
Raw Sample Relinquished by: RM (wc)