

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB136332

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/01/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	07/01/2025	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	07/01/2025	08:34
4	ICV	1	Water	NA	NA	20.2	7.00	07/01/2025	08:37
5	CCV1	1	Water	NA	NA	20.2	2.01	07/01/2025	08:40
6	Q2459-04	1	Solid	20.02	20	21.7	9.70	07/01/2025	09:00
7	Q2459-04DUP	1	Solid	20.03	20	21.8	9.71	07/01/2025	09:04
8	Q2469-04	1	Solid	20.05	20	24.3	9.07	07/01/2025	09:10
9	CCV2	1	Water	NA	NA	20.3	12.02	07/01/2025	09:11

WORKLIST(Hardcopy Internal Chain)

UP 136332

WorkList Name : corrosivity q2469

WorkList ID : 190469

Department : Wet-Chemistry

Date : 07-01-2025 08:16:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2459-04	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	A51	06/28/2025	9045D
Q2469-04	WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	A61	06/30/2025	9045D

Date/Time 07/01/25 08:25
 Raw Sample Received by: SP WCC
 Raw Sample Relinquished by: JDCSM

Date/Time 07/01/25 12:30
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: SP WCC