

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136333

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/01/2025	08:35
2	CAL2	1	Water	NA	NA	20.2	7.00	07/01/2025	08:36
3	CAL3	1	Water	NA	NA	20.2	10.02	07/01/2025	08:39
4	ICV	1	Water	NA	NA	20.3	7.02	07/01/2025	08:40
5	CCV1	1	Water	NA	NA	20.1	2.01	07/01/2025	08:42
6	Q2462-02	1	Solid	20.02	20	22.7	5.09	07/01/2025	09:10
7	Q2462-02DUP	1	Solid	20.03	20	22.9	5.10	07/01/2025	09:13
8	Q2462-03	1	Solid	20.02	20	22.9	6.59	07/01/2025	09:15
9	Q2462-06	1	Solid	20.03	20	22.9	6.00	07/01/2025	09:24
10	Q2462-07	1	Solid	20.04	20	22.8	7.04	07/01/2025	09:30
11	Q2462-08	1	Solid	20.02	20	22.9	6.75	07/01/2025	09:35
12	Q2462-10	1	Solid	20.01	20	21.6	7.08	07/01/2025	09:40
13	CCV2	1	Water	NA	NA	20.2	12.02	07/01/2025	09:42

WORKLIST(Hardcopy Internal Chain)

VB 136233

WorkList Name : ph w q2464

WorkList ID : 190468

Department : Wet-Chemistry

Date : 07-01-2025 08:15:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2462-02	60425-A	Solid	pH	Cool 4 deg C	PSEG03	A12	06/30/2025	9045D
Q2462-03	60425-3	Solid	pH	Cool 4 deg C	PSEG03	A12	06/30/2025	9045D
Q2462-06	61625-A	Solid	pH	Cool 4 deg C	PSEG03	A12	06/30/2025	9045D
Q2462-07	61625-B	Solid	pH	Cool 4 deg C	PSEG03	A12	06/30/2025	9045D
Q2462-08	61625-C	Solid	pH	Cool 4 deg C	PSEG03	A12	06/30/2025	9045D
Q2462-10 C	62725	Solid	pH	Cool 4 deg C	PSEG03	A12	06/30/2025	9045D

Date/Time 07/01/25 08:25

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

Date/Time 07/01/25

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC