

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136334

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/01/2025	08:35
2	CAL2	1	Water	NA	NA	20.2	7.00	07/01/2025	08:36
3	CAL3	1	Water	NA	NA	20.2	10.02	07/01/2025	08:39
4	ICV	1	Water	NA	NA	20.3	7.02	07/01/2025	08:40
5	CCV1	1	Water	NA	NA	20.3	2.02	07/01/2025	09:50
6	Q2462-01	1	Water	NA	NA	26.3	7.70	07/01/2025	10:00
7	Q2462-01DUP	1	Water	NA	NA	26.6	7.72	07/01/2025	10:02
8	Q2465-01	1	Water	NA	NA	26.6	4.79	07/01/2025	10:15
9	Q2465-02	1	Water	NA	NA	27.2	4.93	07/01/2025	10:25
10	Q2465-03	1	Water	NA	NA	27.3	4.83	07/01/2025	10:37
11	CCV2	1	Water	NA	NA	20.3	12.02	07/01/2025	10:40

WORKLIST(Hardcopy Internal Chain)

VB 136334

WorkList Name : ph q2462 s

WorkList ID : 190467

Department : Wet-Chemistry

Date : 07-01-2025 08:14:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2462-01	A 40425	Water	pH	Cool 4 deg C	PSEG03	A12	06/30/2025	9040C
Q2465-01	62025-A	Water	pH	Cool 4 deg C	PSEG03	A42	06/30/2025	9040C
Q2465-02	C 62025-B	Water	pH	Cool 4 deg C	PSEG03		06/30/2025	9040C
Q2465-03	62025-C	Water	pH	Cool 4 deg C	PSEG03	A42	06/30/2025	9040C

Date/Time 07/01/25 08:25

Raw Sample Received by: SP CDC

Raw Sample Relinquished by: SP CDC

Date/Time 07/01/25 13:30

Raw Sample Received by: SP CDC

Raw Sample Relinquished by: SP CDC