

LB136337


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36797



A= 0.0000e+000

B= 2.7123e-004

C= 1.0435e-001

Rho= 0.9993276

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.094	0.094	-38	0.000	-38	0				18
0.2	0.200	0.236	0.036	487	0.0 %	487					0.1
2.5	2.500	2.519	0.019	8901	0.0 %	8901					6
5.0	5.000	4.703	-0.297	16955	0.0 %	16955					1
7.5	7.500	7.563	0.063	27499	0.0 %	27499					
10.0	10.000	10.085	0.085	36797	0.0 %	36797					

LB136337 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Jul 2025 11:11:18

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Std Conc	Date	Method
S	0.0 - 1	50	-	-38	PPB	0.0000	01 Jul 2025 11:19:00	7470A
S	0.2 - 1	50.2	-	0	PPB	0.2000	01 Jul 2025 11:21:17	7470A
S	0.2 - 1	52.5	-	487	PPB	0.2000	01 Jul 2025 11:23:26	7470A
S	2.5 - 1	55	-	8901	PPB	2.5000	01 Jul 2025 11:25:43	7470A
S	5.0 - 1	57.5	-	16955	PPB	5.0000	01 Jul 2025 11:28:00	7470A
S	7.5 - 1	59.5	-	27499	PPB	7.5000	01 Jul 2025 11:30:17	7470A
S	10.0 - 1	514	-	36797	PPB	10.0000	01 Jul 2025 11:32:35	7470A
U	ICV13 - 1	ICV13	3.9656	14236	PPB		01 Jul 2025 11:43:02	7470A
U	ICB13 - 1	ICB13	0.0553	-181	PPB		01 Jul 2025 11:45:17	7470A
U	CCV40 - 1	CCV40	4.5922	16546	PPB		01 Jul 2025 11:47:34	7470A
U	CCB40 - 1	CCB40	0.0436	-224	PPB		01 Jul 2025 11:49:49	7470A
U	CRA - 1	CRA	0.2229	437	PPB		01 Jul 2025 11:52:06	7470A
U	HighStd - 1	HighStd	9.6851	35323	PPB		01 Jul 2025 11:54:22	7470A
U	ChkStd - 1	ChkStd	6.7862	24635	PPB		01 Jul 2025 11:56:37	7470A
U	PB168670BL - 1	PBW	-0.0185	-453	PPB		01 Jul 2025 11:58:56	7470A
U	PB168670BS - 1	LCSW	4.0125	14409	PPB		01 Jul 2025 12:01:14	7470A
U	Q2414-04 - 1	WC-1	0.0431	-226	PPB		01 Jul 2025 12:03:31	7470A
U	Q2415-04 - 1	WC-1	0.1008	-13	PPB		01 Jul 2025 12:05:50	7470A
U	Q2416-04 - 1	MH-G/H	0.1057	5	PPB		01 Jul 2025 12:08:07	7470A
U	Q2420-02 - 1	72-11933	0.1087	16	PPB		01 Jul 2025 12:10:24	7470A
U	Q2429-04 - 1	TP-4	0.1084	15	PPB		01 Jul 2025 12:12:39	7470A
U	CCV41 - 1	CCV41	4.5691	16461	PPB		01 Jul 2025 12:17:39	7470A
U	CCB41 - 1	CCB41	0.0357	-253	PPB		01 Jul 2025 12:19:54	7470A
U	Q2430-04 - 1	MH-E/F	0.1006	-14	PPB		01 Jul 2025 12:22:13	7470A
U	Q2430-04DUP - 1	MH-E/FDUP	0.1046	1	PPB		01 Jul 2025 12:24:28	7470A
U	Q2430-04MS - 1	MH-E/FMS	3.7885	13583	PPB		01 Jul 2025 12:26:43	7470A
U	Q2430-04MSD - 1	MH-E/FMSD	4.0003	14364	PPB		01 Jul 2025 12:28:59	7470A
U	PB168683BL - 1	PBW	0.0544	-184	PPB		01 Jul 2025 12:31:16	7470A
U	PB168683BS - 1	LCSW	3.9708	14255	PPB		01 Jul 2025 12:33:35	7470A
U	Q2458-10 - 1	FB-06272025	0.0547	-183	PPB		01 Jul 2025 12:35:51	7470A
U	Q2463-01 - 1	TW-WTS-11	0.1882	309	PPB		01 Jul 2025 12:38:10	7470A
U	Q2463-01DUP - 1	TW-WTS-11DUP	0.1827	289	PPB		01 Jul 2025 12:40:27	7470A
U	Q2463-01MS - 1	TW-WTS-11MS	3.2371	11550	PPB		01 Jul 2025 12:42:43	7470A
U	CCV42 - 1	CCV42	4.6559	16781	PPB		01 Jul 2025 12:47:41	7470A
U	CCB42 - 1	CCB42	0.0303	-273	PPB		01 Jul 2025 12:49:57	7470A
U	Q2463-01MSD - 1	TW-WTS-11MSD	3.8802	13921	PPB		01 Jul 2025 12:52:15	7470A
U	Q2465-04 - 1	62025-ABC	2.7920	9909	PPB		01 Jul 2025 12:54:31	7470A
U	PB168626TB - 1	PB168626TB	0.0148	-330	PPB		01 Jul 2025 12:56:49	7470A
U	Q2430-04LX5 - 1		0.0989	-20	PPB		01 Jul 2025 12:59:07	7470A
U	Q2430-04A - 1		3.6721	13154	PPB		01 Jul 2025 13:01:22	7470A
U	Q2463-01LX5 - 1		0.0791	-93	PPB		01 Jul 2025 13:03:39	7470A
U	Q2463-01A - 1		3.7790	13548	PPB		01 Jul 2025 13:05:57	7470A
U	CCV43 - 1	CCV43	4.9534	17878	PPB		01 Jul 2025 13:10:53	7470A
U	CCB43 - 1	CCB43	0.0414	-232	PPB		01 Jul 2025 13:13:11	7470A