

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136338

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	07/01/2025	08:45
2	Q2462-01		1	15	100.00	>212.0	>212.0	07/01/2025	09:15
3	Q2462-01DUP		1	15	100.00	>212.0	>212.0	07/01/2025	09:45
4	Q2463-01		1	14	100.00	>212.0	>212.0	07/01/2025	10:15
5	Q2465-04		1	15	100.00	>212.0	>212.0	07/01/2025	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB136338

WorkList Name : fp-7-01

WorkList ID : 190473

Department : Wet-Chemistry

Date : 07-01-2025 08:14:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2462-01	40425	Water	Flash Point	Cool 4 deg C	PSEG03	A12	06/30/2025	1010B
Q2463-01	TW-WTS-11	Water	Flash Point	Cool 4 deg C	ENTA05	A53	06/27/2025	1010B
Q2465-04	62025-ABC	Water	Flash Point	Cool 4 deg C	PSEG03	A42	06/30/2025	1010B

Date/Time 07/01/2025 08.30
 Raw Sample Received by: RY CW
 Raw Sample Relinquished by: SP CW

Date/Time 07/01/2025 11.25
 Raw Sample Received by: SP CW
 Raw Sample Relinquished by: RY CW