

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136349

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/02/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/02/2025	08:02
3	CAL3	1	Water	NA	NA	20.2	10.02	07/02/2025	08:05
4	ICV	1	Water	NA	NA	20.2	7.01	07/02/2025	08:10
5	CCV1	1	Water	NA	NA	20.2	2.02	07/02/2025	08:15
6	Q2477-01	1	Water	NA	NA	24.1	6.40	07/02/2025	08:35
7	Q2477-01DUP	1	Water	NA	NA	24.3	6.42	07/02/2025	08:39
8	CCV2	1	Water	NA	NA	20.2	12.02	07/02/2025	08:42

WORKLIST(Hardcopy Internal Chain)

UB 136349

WorkList Name : ph w q2477

WorkList ID : 190493

Department : Wet-Chemistry

Date : 07-02-2025 07:39:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2477-01	✓ 50728	Water	pH	Cool 4 deg C	PSEG03	A53	07/01/2025	9040C

Date/Time 07/02/25 07:50
Raw Sample Received by: SA COOC
Raw Sample Relinquished by: SA SM

Date/Time 07/02/25 21:00
Raw Sample Received by: SA SM
Raw Sample Relinquished by: SA COOC