

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136350

Slope : 98.3

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/02/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	07/02/2025	08:02
3	CAL3	1	Water	NA	NA	20.2	10.02	07/02/2025	08:05
4	ICV	1	Water	NA	NA	20.2	7.01	07/02/2025	08:10
5	CCV1	1	Water	NA	NA	20.1	2.01	07/02/2025	08:50
6	Q2462-05	1	Solid	20.02	20	22.7	5.61	07/02/2025	09:10
7	Q2462-05DUP	1	Solid	20.04	20	22.9	5.62	07/02/2025	09:15
8	CCV2	1	Water	NA	NA	20.3	12.02	07/02/2025	09:20

WORKLIST(Hardcopy Internal Chain)

VB 136350

WorkList Name : ph q 2462 solids

WorkList ID : 190492

Department : Wet-Chemistry

Date : 07-02-2025 07:38:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2462-05	C 50725	Solid	pH	Cool 4 deg C	PSEG03	A12	06/30/2025	9045D

Date/Time 07/02/25 07:50

Raw Sample Received by: SA CORC

Raw Sample Relinquished by: SA CORC

Date/Time 07/02/25

Raw Sample Received by: SA CORC

Raw Sample Relinquished by: SA CORC